

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE080118\
 Data File : BE097132.D
 Acq On : 1 Aug 2018 16:11
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SICV79

Quant Time: Aug 01 16:50:31 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE080118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 01 16:16:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	1440	0.40	ng/ul	0.00
2) Naphthalene-d8	10.13	136	6920	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.02	164	3742	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.77	188	9903	0.40	ng/ul	0.00
16) Chrysene-d12	20.97	240	12294	0.40	ng/ul	0.00
20) Perylene-d12	23.21	264	11793	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.73	152	4439	0.40	ng/ul	0.00
14) Fluoranthene-d10	18.81	212	12895	0.38	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.18	128	6652	0.397	ng/ul#	93
5) 2-Methylnaphthalene	11.81	142	4306	0.389	ng/ul	98
7) Acenaphthylene	13.74	152	6342	0.392	ng/ul	97
8) Acenaphthene	14.08	153	4637	0.396	ng/ul	95
9) Fluorene	15.08	166	5234	0.389	ng/ul	94
11) Pentachlorophenol	16.44	266	483	0.340	ng/ul	96
12) Phenanthrene	16.81	178	9942	0.394	ng/ul#	89
13) Anthracene	16.90	178	9157	0.389	ng/ul#	91
15) Fluoranthene	18.84	202	12688	0.380	ng/ul	84
17) Pyrene	19.21	202	13326	0.405	ng/ul#	78
18) Benzo(a)anthracene	20.96	228	13786	0.392	ng/ul#	86
19) Chrysene	21.01	228	14329	0.396	ng/ul	97
21) Benzo(b)fluoranthene	22.54	252	13086	0.407	ng/ul	84
22) Benzo(k)fluoranthene	22.58	252	13732	0.396	ng/ul#	86
23) Benzo(a)pyrene	23.11	252	12534	0.390	ng/ul#	78
24) Indeno(1,2,3-cd)pyrene	25.50	276	15841	0.394	ng/ul#	77
25) Dibenzo(a,h)anthracene	25.53	278	13012	0.395	ng/ul#	88
26) Benzo(g,h,i)perylene	26.20	276	14062	0.404	ng/ul#	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE080118\
Data File : BE097132.D
Acq On : 1 Aug 2018 16:11
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SICV79

Quant Time: Aug 01 16:50:31 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE080118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 01 16:16:30 2018
Response via : Initial Calibration

