

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE080318\
 Data File : BE097188.D
 Acq On : 3 Aug 2018 20:32
 Operator : SJ/JU
 Sample : J4171-07
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 JJ1J9

Manual Integrations
 APPROVED

Sohil
 8/6/2018 4:32:32 PM

Quant Time: Aug 04 00:20:07 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE080118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 03 23:39:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.39	152	1684	0.40	ng/ul	0.00
2) Naphthalene-d8	10.14	136	8823	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.02	164	5784	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.78	188	12695m	0.40	ng/ul	0.00
16) Chrysene-d12	20.98	240	13601m	0.40	ng/ul	0.00
20) Perylene-d12	23.22	264	13550m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.73	152	4276	0.30	ng/ul	0.00
14) Fluoranthene-d10	18.81	212	11831m	0.27	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.18	128	1253	0.059	ng/ul#	89
5) 2-Methylnaphthalene	11.81	142	626	0.044	ng/ul	96
7) Acenaphthylene	13.74	152	1447	0.058	ng/ul#	81
12) Phenanthrene	16.81	178	7060m	0.218	ng/ul	
13) Anthracene	16.91	178	1631m	0.054	ng/ul	
15) Fluoranthene	18.84	202	12629m	0.295	ng/ul	
17) Pyrene	19.21	202	17242	0.473	ng/ul#	84
18) Benzo(a)anthracene	20.96	228	8362m	0.215	ng/ul	
19) Chrysene	21.01	228	10025m	0.250	ng/ul	
21) Benzo(b)fluoranthene	22.55	252	14036m	0.380	ng/ul	
22) Benzo(k)fluoranthene	22.57	252	5046m	0.127	ng/ul	
23) Benzo(a)pyrene	23.13	252	11084m	0.300	ng/ul	
24) Indeno(1,2,3-cd)pyrene	25.51	276	9137m	0.198	ng/ul	
25) Dibenzo(a,h)anthracene	25.53	278	2851m	0.075	ng/ul	
26) Benzo(g,h,i)perylene	26.22	276	10376m	0.260	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE080318\
Data File : BE097188.D
Acq On : 3 Aug 2018 20:32
Operator : SJ/JU
Sample : J4171-07
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
JJ1J9

Manual Integrations
APPROVED

Sohil
8/6/2018 4:32:32 PM

Quant Time: Aug 04 00:20:07 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE080118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 03 23:39:09 2018
Response via : Initial Calibration

