

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE080418\
 Data File : BE097218.D
 Acq On : 5 Aug 2018 2:11
 Operator : SJ/JU
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD0.490

Quant Time: Aug 06 03:52:25 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE080118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 06 03:46:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.39	152	1522	0.40	ng/ul	0.00
2) Naphthalene-d8	10.13	136	8279	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.02	164	4849	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.77	188	11346	0.40	ng/ul	0.00
16) Chrysene-d12	20.97	240	11817	0.40	ng/ul	0.00
20) Perylene-d12	23.21	264	12141	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.72	152	5430	0.41	ng/ul	0.00
14) Fluoranthene-d10	18.81	212	13301	0.34	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.18	128	7723	0.385	ng/ul#	92
5) 2-Methylnaphthalene	11.80	142	5390	0.407	ng/ul	98
7) Acenaphthylene	13.73	152	8305	0.396	ng/ul	97
8) Acenaphthene	14.08	153	5899	0.389	ng/ul	93
9) Fluorene	15.08	166	6746	0.387	ng/ul#	93
11) Pentachlorophenol	16.44	266	569	0.350	ng/ul	96
12) Phenanthrene	16.81	178	11272	0.390	ng/ul#	90
13) Anthracene	16.90	178	10886	0.403	ng/ul#	92
15) Fluoranthene	18.84	202	13167	0.344	ng/ul	84
17) Pyrene	19.20	202	13227	0.418	ng/ul#	80
18) Benzo(a)anthracene	20.96	228	13606	0.403	ng/ul#	84
19) Chrysene	21.01	228	13650	0.392	ng/ul	97
21) Benzo(b)fluoranthene	22.54	252	13453	0.407	ng/ul	84
22) Benzo(k)fluoranthene	22.58	252	13912	0.390	ng/ul#	88
23) Benzo(a)pyrene	23.11	252	13382	0.404	ng/ul#	78
24) Indeno(1,2,3-cd)pyrene	25.50	276	16867	0.407	ng/ul#	77
25) Dibenzo(a,h)anthracene	25.53	278	13968	0.412	ng/ul#	89
26) Benzo(g,h,i)perylene	26.20	276	14206	0.397	ng/ul#	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE080418\
Data File : BE097218.D
Acq On : 5 Aug 2018 2:11
Operator : SJ/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD0.490

Quant Time: Aug 06 03:52:25 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE080118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 06 03:46:27 2018
Response via : Initial Calibration

