

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE080618\
 Data File : BE097224.D
 Acq On : 6 Aug 2018 14:53
 Operator : SJ/JU
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD0.492

Quant Time: Aug 06 17:46:51 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE080118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 06 17:09:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.39	152	1247	0.40	ng/ul	0.00
2) Naphthalene-d8	10.13	136	6868	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.02	164	4020	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.77	188	9447	0.40	ng/ul	0.00
16) Chrysene-d12	20.97	240	9444	0.40	ng/ul	0.00
20) Perylene-d12	23.21	264	9572	0.40	ng/ul	0.00

System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.72	152	4517	0.41	ng/ul	0.00
14) Fluoranthene-d10	18.81	212	10612	0.33	ng/ul	0.00

Target Compounds				Qvalue		
3) Naphthalene	10.18	128	6452	0.388	ng/ul#	92
5) 2-Methylnaphthalene	11.80	142	4460	0.406	ng/ul	98
7) Acenaphthylene	13.73	152	6929	0.399	ng/ul	97
8) Acenaphthene	14.08	153	4938	0.393	ng/ul	94
9) Fluorene	15.08	166	5574	0.386	ng/ul#	94
11) Pentachlorophenol	16.44	266	492	0.363	ng/ul	98
12) Phenanthrene	16.81	178	9251	0.384	ng/ul#	89
13) Anthracene	16.90	178	8883	0.395	ng/ul#	93
15) Fluoranthene	18.84	202	10466	0.329	ng/ul	84
17) Pyrene	19.20	202	10327	0.408	ng/ul#	80
18) Benzo(a)anthracene	20.96	228	10571	0.392	ng/ul#	85
19) Chrysene	21.01	228	10785	0.388	ng/ul	97
21) Benzo(b)fluoranthene	22.53	252	10351	0.397	ng/ul	84
22) Benzo(k)fluoranthene	22.58	252	10851	0.386	ng/ul#	88
23) Benzo(a)pyrene	23.11	252	10409	0.399	ng/ul#	79
24) Indeno(1,2,3-cd)pyrene	25.50	276	13004	0.398	ng/ul#	76
25) Dibenzo(a,h)anthracene	25.53	278	10780	0.403	ng/ul#	89
26) Benzo(g,h,i)perylene	26.21	276	10986	0.389	ng/ul#	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE080618\
Data File : BE097224.D
Acq On : 6 Aug 2018 14:53
Operator : SJ/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD0.492

Quant Time: Aug 06 17:46:51 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE080118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 06 17:09:49 2018
Response via : Initial Calibration

