

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE081018\
 Data File : BE097258.D
 Acq On : 10 Aug 2018 12:12
 Operator : SJ/JU
 Sample : J4379-12
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MW-103S

Quant Time: Aug 10 16:29:08 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 31 10:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.39	152	1097	0.40	ng	0.00
7) Naphthalene-d8	10.15	136	5739	0.40	ng	0.00
13) Acenaphthene-d10	14.01	164	3628	0.40	ng	0.00
19) Phenanthrene-d10	16.77	188	9101	0.40	ng	0.00
27) Chrysene-d12	20.98	240	9879	0.40	ng	0.00
34) Perylene-d12	23.22	264	9351	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.05	112	659	0.21	ng	0.00
5) Phenol-d6	6.59	99	529	0.12	ng	0.00
8) Nitrobenzene-d5	8.52	82	2012	0.40	ng	0.00
11) 2-Methylnaphthalene-d10	11.72	152	3503	0.41	ng	0.00
14) 2,4,6-Tribromophenol	15.52	330	705	0.59	ng	0.00
15) 2-Fluorobiphenyl	12.63	172	5833	0.47	ng	0.00
25) Fluoranthene-d10	18.81	212	40705	0.45	ng	0.00
29) Terphenyl-d14	19.43	244	7571	0.40	ng	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.09	88	1119	0.642	ng	92
9) Naphthalene	10.19	128	297661	5.780	ng	99
12) 2-Methylnaphthalene	11.80	142	4845	0.548	ng	98
16) Acenaphthylene	13.73	152	1813	0.111	ng	# 62
17) Acenaphthene	14.07	154	62006	6.550	ng	95
18) Fluorene	15.06	166	24553	1.963	ng	# 80
23) Phenanthrene	16.80	178	53654	2.473	ng	97
24) Anthracene	16.89	178	14780	0.762	ng	# 87
26) Fluoranthene	18.84	202	95489	3.923	ng	98
28) Pyrene	19.20	202	62816	2.318	ng	# 91
30) Benzo(a)anthracene	20.95	228	7100	0.289	ng	# 75
31) Chrysene	21.02	228	5954	0.226	ng	# 61
32) Bis(2-ethylhexyl)phthalate	20.92	149	29272	1.028	ng	# 97
35) Benzo(b)fluoranthene	22.54	252	1697	0.072	ng	# 30

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE081018\
Data File : BE097258.D
Acq On : 10 Aug 2018 12:12
Operator : SJ/JU
Sample : J4379-12
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MW-103S

Quant Time: Aug 10 16:29:08 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 31 10:49:28 2018
Response via : Initial Calibration

