

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE081018\
 Data File : BE097264.D
 Acq On : 10 Aug 2018 15:59
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 10 16:49:40 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 31 10:49:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2	1,4-Dioxane	0.636	0.460	27.7#	62	0.00
3	n-Nitrosodimethylamine	0.711	0.513	27.8#	65	0.00
4 S	2-Fluorophenol	1.124	1.233	-9.7	97	0.01
5 S	Phenol-d6	1.649	1.900	-15.2	103	0.01
6	bis(2-Chloroethyl)ether	1.506	1.423	5.5	85	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
8 S	Nitrobenzene-d5	0.347	0.361	-4.0	100	0.00
9	Naphthalene	3.590	3.482	3.0	93	0.00
10	Hexachlorobutadiene	0.161	0.151	6.2	89	0.00
11 SURR	2-Methylnaphthalene-d10	0.599	0.554	7.5	88	0.00
12	2-Methylnaphthalene	0.616	0.578	6.2	89	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	84	0.00
14 S	2,4,6-Tribromophenol	0.132	0.139	-5.3	98	0.01
15 S	2-Fluorobiphenyl	1.381	1.344	2.7	84	0.00
16	Acenaphthylene	1.795	1.712	4.6	83	0.00
17	Acenaphthene	1.044	1.001	4.1	83	0.00
18	Fluorene	1.379	1.233	10.6	79	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	70	0.00
20	4-Bromophenyl-phenylether	0.194	0.196	-1.0	74	0.00
21	Hexachlorobenzene	0.209	0.217	-3.8	74	0.00
22	Pentachlorophenol	0.057	0.064	-12.3	99	0.01
23	Phenanthrene	0.953	0.908	4.7	70	0.00
24	Anthracene	0.853	0.836	2.0	74	0.01
25 SURR	Fluoranthene-d10	4.003	3.924	2.0	75	0.00
26	Fluoranthene	1.070	1.030	3.7	72	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	80	0.00
28	Pyrene	1.097	1.010	7.9	77	0.00
29 S	Terphenyl-d14	0.768	0.708	7.8	79	0.00
30	Benzo(a)anthracene	0.995	1.044	-4.9	93	0.00
31	Chrysene	1.067	0.972	8.9	75	0.01
32	Bis(2-ethylhexyl)phthalate	1.243	2.197	-76.7#	173#	0.00
33	Indeno(1,2,3-cd)pyrene	0.966	1.257	-30.1#	112	0.01
34 I	Perylene-d12	1.000	1.000	0.0	100	0.00
35	Benzo(b)fluoranthene	1.015	0.939	7.5	98	0.00
36	Benzo(k)fluoranthene	1.185	1.077	9.1	98	0.00
37 C	Benzo(a)pyrene	0.955	0.960	-0.5	106	0.00
38	Dibenzo(a,h)anthracene	0.953	1.015	-6.5	112	0.01
39	Benzo(g,h,i)perylene	1.011	0.984	2.7	102	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE081018\
Data File : BE097264.D
Acq On : 10 Aug 2018 15:59
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 10 16:49:40 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 31 10:49:28 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		