

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE081018\
 Data File : BE097264.D
 Acq On : 10 Aug 2018 15:59
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

Sohil
 8/10/2018 4:52:12 PM

Quant Time: Aug 10 16:49:40 2018

Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072418.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Tue Jul 31 10:49:28 2018

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.39	152	1153	0.40	ng	0.00
7) Naphthalene-d8	10.14	136	6266	0.40	ng	0.00
13) Acenaphthene-d10	14.01	164	3649	0.40	ng	0.00
19) Phenanthrene-d10	16.77	188	8147	0.40	ng	0.00
27) Chrysene-d12	20.98	240	8597	0.40	ng	0.00
34) Perylene-d12	23.22	264	8906	0.40	ng	0.00

System Monitoring Compounds

4) 2-Fluorophenol	5.07	112	1422	0.44	ng	0.01
5) Phenol-d6	6.61	99	2191	0.46	ng	0.01
8) Nitrobenzene-d5	8.52	82	2265	0.42	ng	0.00
11) 2-Methylnaphthalene-d10	11.73	152	3471	0.37	ng	0.00
14) 2,4,6-Tribromophenol	15.53	330	506	0.42	ng	0.01
15) 2-Fluorobiphenyl	12.63	172	4905	0.39	ng	0.00
25) Fluoranthene-d10	18.81	212	31972	0.39	ng	0.00
29) Terphenyl-d14	19.43	244	6084	0.37	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.10	88	530	0.289	ng	# 91
3) n-Nitrosodimethylamine	3.38	42	591	0.288	ng	# 88
6) bis(2-Chloroethyl)ether	6.84	93	1641	0.378	ng	93
9) Naphthalene	10.20	128	21816	0.388	ng	100
10) Hexachlorobutadiene	10.48	225	948	0.376	ng	96
12) 2-Methylnaphthalene	11.80	142	3619	0.375	ng	99
16) Acenaphthylene	13.73	152	6247	0.381	ng	99
17) Acenaphthene	14.08	154	3653	0.384	ng	94
18) Fluorene	15.08	166	4498	0.357	ng	# 79
20) 4-Bromophenyl-phenylether	15.97	248	1594	0.403	ng	87
21) Hexachlorobenzene	16.08	284	1767	0.416	ng	# 84
22) Pentachlorophenol	16.44	266	524	0.508	ng	# 78
23) Phenanthrene	16.81	178	7401	0.381	ng	98
24) Anthracene	16.90	178	6813	0.392	ng	96
26) Fluoranthene	18.84	202	8389	0.385	ng	98
28) Pyrene	19.21	202	8681	0.368	ng	99
30) Benzo(a)anthracene	20.96	228	8974	0.420	ng	98
31) Chrysene	21.02	228	8360	0.364	ng	99
32) Bis(2-ethylhexyl)phthalate	20.92	149	18886	0.792	ng	99
33) Indeno(1,2,3-cd)pyrene	25.52	276	10807	0.520	ng	# 82
35) Benzo(b)fluoranthene	22.54	252	8361m	0.370	ng	
36) Benzo(k)fluoranthene	22.58	252	9595	0.364	ng	# 86
37) Benzo(a)pyrene	23.12	252	8551	0.402	ng	# 92
38) Dibenzo(a,h)anthracene	25.54	278	9036	0.426	ng	95
39) Benzo(g,h,i)perylene	26.22	276	8762	0.389	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE081018\
Data File : BE097264.D
Acq On : 10 Aug 2018 15:59
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

Sohil
8/10/2018 4:52:12 PM

Quant Time: Aug 10 16:49:40 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 31 10:49:28 2018
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.39 min Scan# 393
Delta R.T. 0.00 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
152	100		
150	148.5	117.2	175.8
115	66.5	57.0	85.6

Manual Integrations
APPROVED

Sohil
8/10/2018 4:52:12 PM






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane
Concen: 0.289 ng
RT: 3.10 min Scan# 20
Delta R.T. 0.00 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Tgt Ion	Ratio	Lower	Upper
88	100		
58	80.6	63.2	94.8
43	37.0	41.2	61.8#






Abundance

Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0









#3
n-Nitrosodimethylamine
Concen: 0.288 ng
RT: 3.38 min Scan# 45
Delta R.T. 0.00 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
42	100		
74	129.4	96.0	144.0
44	0.0	12.0	18.0




Manual Integrations
APPROVED




Sohil
8/10/2018 4:52:12 PM



#4
2-Fluorophenol
Concen: 0.439 ng
RT: 5.07 min Scan# 191
Delta R.T. 0.01 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Tgt Ion	Ratio	Lower	Upper
112	100		
64	67.0	60.1	90.1
63	141.9	116.8	175.2





#5

Phenol-d6

Concen: 0.461 ng

RT: 6.61 min Scan# 325

Delta R.T. 0.01 min

Lab File: BE097264.D

Acq: 10 Aug 2018 15:59

Tgt Ion:	99	Resp:	2191
Ion	Ratio	Lower	Upper
99	100		
42	21.4	20.2	30.2
71	35.3	28.4	42.6

Instrument :





BNA_E

ClientSampleId :

SSTDCCC0.4EC

Manual Integrations
APPROVED

Sohil
8/10/2018 4:52:12 PM



#6

bis(2-Chloroethyl)ether

Concen: 0.378 ng


RT: 6.84 min Scan# 345


Delta R.T. 0.00 min

Lab File: BE097264.D

Acq: 10 Aug 2018 15:59

Tgt Ion:	93	Resp:	1641
Ion	Ratio	Lower	Upper
93	100		
63	328.0	275.3	412.9
95	31.4	25.2	37.8





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.14 min Scan# 613

Delta R.T. -0.00 min

Lab File: BE097264.D

Acq: 10 Aug 2018 15:59

Instrument :

BNA_E


ClientSampleId :

SSTDCCC0.4EC




Tgt Ion	Ratio	Lower	Upper
136	100		
137	12.1	9.5	14.3
54	34.0	29.1	43.7
68	6.8	6.2	9.2

Manual Integrations
APPROVED

Sohil
8/10/2018 4:52:12 PM



Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BE0
Ion 68.00 (67.70 to 68.70): BE0



#8

Nitrobenzene-d5

Concen: 0.416 ng


RT: 8.52 min Scan# 491

Delta R.T. 0.00 min



Lab File: BE097264.D


Acq: 10 Aug 2018 15:59

Tgt Ion	Ratio	Lower	Upper
82	100		
128	30.0	24.0	36.0
54	57.4	40.0	60.0



Abundance Ion 82.00 (81.70 to 82.70): BE0
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BE0





#9

Naphthalene

Concen: 0.388 ng

RT: 10.20 min Scan# 617

Delta R.T. -0.00 min

Lab File: BE097264.D

Acq: 10 Aug 2018 15:59

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	128	Resp:	21816
Ion Ratio	Lower	Upper	
128	100		
129	3.1	2.6	3.8
127	3.4	2.8	4.2

Manual Integrations
APPROVED

Sohil
8/10/2018 4:52:12 PM



#10

Hexachlorobutadiene

Concen: 0.376 ng




RT: 10.48 min Scan# 639


Delta R.T. -0.00 min

Lab File: BE097264.D

Acq: 10 Aug 2018 15:59

Tgt Ion:	225	Resp:	948
Ion Ratio	Lower	Upper	
225	100		
223	63.2	53.0	79.6
227	66.9	51.4	77.2









#11
 2-Methylnaphthalene-d10
 Concen: 0.370 ng
 RT: 11.73 min Scan# 761
 Delta R.T. 0.00 min
 Lab File: BE097264.D
 Acq: 10 Aug 2018 15:59

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.0	15.4	23.0




Manual Integrations
 APPROVED


Sohil
 8/10/2018 4:52:12 PM



#12
 2-Methylnaphthalene
 Concen: 0.375 ng
 RT: 11.80 min Scan# 777
 Delta R.T. -0.00 min
 Lab File: BE097264.D
 Acq: 10 Aug 2018 15:59

Tgt Ion	Ratio	Lower	Upper
142	100		
141	86.9	70.4	105.6
115	39.6	31.7	47.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.01 min Scan# 1028
Delta R.T. 0.00 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
164	100		
162	103.7	83.1	124.7
160	52.1	37.1	55.7




Manual Integrations
APPROVED


Sohil
8/10/2018 4:52:12 PM



#14
2,4,6-Tribromophenol
Concen: 0.421 ng
RT: 15.53 min Scan# 1155
Delta R.T. 0.01 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Tgt Ion	Ratio	Lower	Upper
330	100		
332	97.6	152.7	229.1#
141	39.7	33.7	50.5









#15
2-Fluorobiphenyl
Concen: 0.389 ng
RT: 12.63 min Scan# 908
Delta R.T. 0.00 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.6	29.9	44.9
170	24.1	21.8	32.8




Manual Integrations
APPROVED


Sohil
8/10/2018 4:52:12 PM



#16
Acenaphthylene
Concen: 0.381 ng
RT: 13.73 min Scan# 1003
Delta R.T. 0.00 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.0	15.7	23.5
153	13.0	10.5	15.7









#17
Acenaphthene
Concen: 0.384 ng
RT: 14.08 min Scan# 1034
Delta R.T. 0.00 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
154	100		
153	114.5	89.2	133.8
152	60.7	42.0	63.0




Manual Integrations
APPROVED


Sohil
8/10/2018 4:52:12 PM



#18
Fluorene
Concen: 0.357 ng
RT: 15.08 min Scan# 1116
Delta R.T. -0.00 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Tgt Ion	Ratio	Lower	Upper
166	100		
165	99.7	77.8	116.6
167	14.0	42.4	63.6#





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.77 min Scan# 1261

Delta R.T. -0.00 min

Lab File: BE097264.D

Acq: 10 Aug 2018 15:59

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	188	Resp:	8147
Ion Ratio	Lower	Upper	
188	100		
94	0.0	3.9	5.9#
80	13.7	3.6	5.4#

Manual Integrations
APPROVED

Sohil
8/10/2018 4:52:12 PM



#20

4-Bromophenyl-phenylether

Concen: 0.403 ng




RT: 15.97 min Scan# 1193


Delta R.T. -0.00 min

Lab File: BE097264.D

Acq: 10 Aug 2018 15:59

Tgt Ion:	248	Resp:	1594
Ion Ratio	Lower	Upper	
248	100		
250	92.6	62.7	94.1
141	67.4	48.2	72.2






#21
Hexachlorobenzene
Concen: 0.416 ng
RT: 16.08 min Scan# 1202
Delta R.T. -0.00 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion	Ratio	Lower	Upper
284	100		
142	32.7	22.1	33.1
249	30.7	34.8	52.2#

Manual Integrations
APPROVED

Sohil
8/10/2018 4:52:12 PM




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#22
Pentachlorophenol
Concen: 0.508 ng
RT: 16.44 min Scan# 1233
Delta R.T. 0.01 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Tgt Ion	Ratio	Lower	Upper
266	100		
264	71.9	72.5	108.7#
268	69.9	73.8	110.6#



Abundance Ion 266.00 (265.70 to 266.70): E
Ion 264.00 (263.70 to 264.70): E
Ion 268.00 (267.70 to 268.70): E









#23
Phenanthrene
Concen: 0.381 ng
RT: 16.81 min Scan# 1265
Delta R.T. -0.00 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.6	15.1	22.7
179	15.6	11.8	17.6




Manual Integrations
APPROVED


Sohil
8/10/2018 4:52:12 PM



#24
Anthracene
Concen: 0.392 ng
RT: 16.90 min Scan# 1273
Delta R.T. 0.01 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.5	12.8	19.2
179	15.5	13.0	19.6





#25

Fluoranthene-d10

Concen: 0.392 ng

RT: 18.81 min Scan# 1503

Delta R.T. 0.00 min

Lab File: BE097264.D

Acq: 10 Aug 2018 15:59

Instrument :

BNA_E


ClientSampleId :

SSTDCCC0.4EC




Tgt Ion:	212	Resp:	31972
Ion Ratio	Lower	Upper	
212	100		
106	3.1	2.8	4.2
104	1.8	1.6	2.4

Manual Integrations
APPROVED

Sohil
8/10/2018 4:52:12 PM



Abundance Ion 212.00 (211.70 to 212.70): E
Ion 106.00 (105.70 to 106.70): E
Ion 104.00 (103.70 to 104.70): E



#26

Fluoranthene

Concen: 0.385 ng


RT: 18.84 min Scan# 1508

Delta R.T. 0.00 min



Lab File: BE097264.D


Acq: 10 Aug 2018 15:59

Tgt Ion:	202	Resp:	8389
Ion Ratio	Lower	Upper	
202	100		
101	10.9	9.8	14.8
203	17.1	13.7	20.5



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E









#27
 Chrysene-d12
 Concen: 0.400 ng
 RT: 20.98 min Scan# 1788
 Delta R.T. 0.00 min
 Lab File: BE097264.D
 Acq: 10 Aug 2018 15:59

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
240	100		
120	6.0	5.5	8.3
236	27.4	20.7	31.1




Manual Integrations
 APPROVED


Sohil
 8/10/2018 4:52:12 PM



#28
 Pyrene
 Concen: 0.368 ng
 RT: 19.21 min Scan# 1572
 Delta R.T. 0.01 min
 Lab File: BE097264.D
 Acq: 10 Aug 2018 15:59

Tgt Ion	Ratio	Lower	Upper
202	100		
200	21.2	16.6	25.0
203	17.6	13.8	20.8









#29
 Terphenyl-d14
 Concen: 0.368 ng
 RT: 19.43 min Scan# 1611
 Delta R.T. 0.00 min
 Lab File: BE097264.D
 Acq: 10 Aug 2018 15:59

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
244	100		
212	33.2	25.4	38.0
122	10.6	8.1	12.1




Manual Integrations
 APPROVED


Sohil
 8/10/2018 4:52:12 PM



#30
 Benzo(a)anthracene
 Concen: 0.420 ng
 RT: 20.96 min Scan# 1786
 Delta R.T. 0.00 min
 Lab File: BE097264.D
 Acq: 10 Aug 2018 15:59

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.6	24.0	36.0
229	19.9	16.0	24.0





#31

Chrysene

Concen: 0.364 ng

RT: 21.02 min Scan# 1791

Delta R.T. 0.01 min

Lab File: BE097264.D

Acq: 10 Aug 2018 15:59

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	228	Resp:	8360
Ion	Ratio	Lower	Upper
228	100		
226	29.9	24.0	36.0
229	20.9	16.0	24.0

Manual Integrations
APPROVED

Sohil
8/10/2018 4:52:12 PM



#32

Bis(2-ethylhexyl)phthalate

Concen: 0.792 ng




RT: 20.92 min Scan# 1783


Delta R.T. 0.00 min

Lab File: BE097264.D

Acq: 10 Aug 2018 15:59

Tgt Ion:	149	Resp:	18886
Ion	Ratio	Lower	Upper
149	100		
167	6.9	5.4	8.0
279	1.1	0.7	1.1






#33
 Indeno(1,2,3-cd)pyrene
 Concen: 0.520 ng
 RT: 25.52 min Scan# 2848
 Delta R.T. 0.01 min
 Lab File: BE097264.D
 Acq: 10 Aug 2018 15:59

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
276	100		
138	10.0	22.5	33.7#
277	24.6	19.9	29.9

Manual Integrations
 APPROVED

Sohil
 8/10/2018 4:52:12 PM






Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#34
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.22 min Scan# 2207
 Delta R.T. 0.01 min
 Lab File: BE097264.D
 Acq: 10 Aug 2018 15:59




Tgt Ion	Ratio	Lower	Upper
264	100		
260	26.2	20.5	30.7
265	31.3	22.8	34.2


Abundance

Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E









#35
Benzo(b)fluoranthene
Concen: 0.370 ng m
RT: 22.54 min Scan# 2017
Delta R.T. 0.00 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.5	20.0	30.0
125	11.3	16.0	24.0




Manual Integrations
APPROVED


Sohil
8/10/2018 4:52:12 PM



#36
Benzo(k)fluoranthene
Concen: 0.364 ng
RT: 22.58 min Scan# 2029
Delta R.T. 0.00 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Tgt Ion	Ratio	Lower	Upper
252	100		
253	25.0	16.0	24.0
125	17.0	8.0	12.0









#37
Benzo(a)pyrene
Concen: 0.402 ng
RT: 23.12 min Scan# 2179
Delta R.T. 0.01 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.8	16.0	24.0#
125	13.1	12.0	18.0




Manual Integrations
APPROVED




Sohil
8/10/2018 4:52:12 PM



#38
Dibenzo(a,h)anthracene
Concen: 0.426 ng
RT: 25.54 min Scan# 2855
Delta R.T. 0.01 min
Lab File: BE097264.D
Acq: 10 Aug 2018 15:59

Tgt Ion	Ratio	Lower	Upper
278	100		
139	22.6	16.0	24.0
279	27.1	20.0	30.0





#39

Benzo(g,h,i)perylene

Concen: 0.389 ng

RT: 26.22 min Scan# 3045

Delta R.T. 0.02 min

Lab File: BE097264.D

Acq: 10 Aug 2018 15:59

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
276	100		
277	26.0	16.0	24.0
138	21.9	20.0	30.0

Manual Integrations
APPROVED

Sohil
8/10/2018 4:52:12 PM

