

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE081518\
 Data File : BE097279.D
 Acq On : 15 Aug 2018 15:04
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICSVBE081518

Quant Time: Aug 15 17:29:46 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE081518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 15 14:23:39 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	99	0.00
2	1,4-Dioxane	0.400	0.381	4.8	99	0.00
3	n-Nitrosodimethylamine	0.400	0.377	5.8	115	0.00
4 S	2-Fluorophenol	0.400	0.369	7.8	99	-0.01
5 S	Phenol-d6	0.400	0.368	8.0	96	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.366	8.5	97	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	96	0.00
8 S	Nitrobenzene-d5	0.400	0.375	6.3	97	0.00
9	Naphthalene	0.400	0.376	6.0	96	-0.01
10	Hexachlorobutadiene	0.400	0.380	5.0	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.376	6.0	95	0.00
12	2-Methylnaphthalene	0.400	0.371	7.3	94	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.400	0.395	1.3	93	0.00
15 S	2-Fluorobiphenyl	0.400	0.385	3.8	95	0.00
16	Acenaphthylene	0.400	0.390	2.5	95	0.00
17	Acenaphthene	0.400	0.393	1.8	95	0.00
18	Fluorene	0.400	0.380	5.0	94	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	96	0.00
20	4-Bromophenyl-phenylether	0.400	0.382	4.5	95	0.00
21	Hexachlorobenzene	0.400	0.392	2.0	96	0.00
22	Pentachlorophenol	0.400	0.360	10.0	92	0.00
23	Phenanthrene	0.400	0.379	5.3	94	0.00
24	Anthracene	0.400	0.361	9.8	92	0.00
25 SURR	Fluoranthene-d10	0.400	0.354	11.5	93	0.00
26	Fluoranthene	0.400	0.358	10.5	93	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	92	0.00
28	Pvrene	0.400	0.382	4.5	91	0.00
29 S	Terphenyl-d14	0.400	0.383	4.3	91	0.00
30	Benzo(a)anthracene	0.400	0.372	7.0	88	0.00
31	Chrysene	0.400	0.395	1.3	94	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.300	25.0#	75	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.370	7.5	87	0.00
34 I	Perylene-d12	0.400	0.400	0.0	88	0.00
35	Benzo(b)fluoranthene	0.400	0.377	5.8	87	0.00
36	Benzo(k)fluoranthene	0.400	0.401	-0.3	91	0.00
37 C	Benzo(a)pyrene	0.400	0.375	6.3	85	0.00
38	Dibenzo(a,h)anthracene	0.400	0.362	9.5	84	0.00
39	Benzo(a,h,i)perylene	0.400	0.386	3.5	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE081518\
Data File : BE097279.D
Acq On : 15 Aug 2018 15:04
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE081518

Quant Time: Aug 15 17:29:46 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE081518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 15 14:23:39 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				