


Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE081518\
 Data File : BE097272.D
 Acq On : 15 Aug 2018 9:52
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE081518.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Aug 15 14:23:39 2018



AutoFind: Scans 2494, 2495, 2496; Background Corrected with Scan 2487

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	50.0	9139	PASS
68	69	0.00	2	1.8	149	PASS
69	198	0.00	100	46.4	8481	PASS
70	69	0.00	2	0.6	47	PASS
127	198	10	80	50.7	9258	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	18276	PASS
199	198	5	9	6.8	1251	PASS
275	198	10	60	28.9	5287	PASS
365	198	1	100	4.0	726	PASS
441	198	0.01	100	15.4	2809	PASS
442	198	50	100	92.2	16845	PASS
443	442	15	24	18.8	3174	PASS