

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE082018\
 Data File : BE097337.D
 Acq On : 20 Aug 2018 18:11
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE082018

Quant Time: Aug 20 18:51:11 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE082018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 20 18:32:10 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.634	0.625	1.4	129	0.00
3	n-Nitrosodimethylamine	0.553	0.639	-15.6	149	0.00
4 S	2-Fluorophenol	0.917	0.986	-7.5	139	0.00
5 S	Phenol-d6	1.359	1.483	-9.1	138	0.00
6	bis(2-Chloroethyl)ether	1.353	1.422	-5.1	129	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	115	0.00
8 S	Nitrobenzene-d5	0.171	0.205	-19.9	153#	0.00
9	Naphthalene	3.666	3.758	-2.5	127	0.00
10	Hexachlorobutadiene	0.161	0.157	2.5	119	-0.01
11 SURR	2-Methylnaphthalene-d10	0.541	0.564	-4.3	132	0.00
12	2-Methylnaphthalene	0.552	0.577	-4.5	133	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	124	0.00
14 S	2,4,6-Tribromophenol	0.066	0.073	-10.6	174#	0.00
15 S	2-Fluorobiphenyl	1.640	1.606	2.1	129	0.00
16	Acenaphthylene	1.522	1.429	6.1	135	0.00
17	Acenaphthene	1.076	1.088	-1.1	136	0.00
18	Fluorene	1.314	1.280	2.6	133	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	124	0.00
20	4-Bromophenyl-phenylether	0.181	0.176	2.8	137	0.00
21	Hexachlorobenzene	0.242	0.233	3.7	125	0.00
22	Pentachlorophenol	0.024	0.020	16.7	170#	-0.01
23	Phenanthrene	0.976	0.984	-0.8	135	0.00
24	Anthracene	0.722	0.756	-4.7	143	0.00
25 SURR	Fluoranthene-d10	3.569	3.522	1.3	140	0.00
26	Fluoranthene	0.940	0.972	-3.4	138	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	130	0.00
28	Pyrene	1.074	1.176	-9.5	152#	0.00
29 S	Terphenyl-d14	0.801	0.843	-5.2	148	0.00
30	Benzo(a)anthracene	0.860	0.895	-4.1	155#	0.00
31	Chrysene	1.127	1.142	-1.3	140	0.00
32	Bis(2-ethylhexyl)phthalate	0.524	0.546	-4.2	157#	0.00
33	Indeno(1,2,3-cd)pyrene	0.716	0.722	-0.8	147	0.00
34 I	Perylene-d12	1.000	1.000	0.0	137	0.00
35	Benzo(b)fluoranthene	1.055	0.993	5.9	144	0.00
36	Benzo(k)fluoranthene	1.480	1.434	3.1	144	0.00
37 C	Benzo(a)pyrene	0.960	1.023	-6.6	162#	0.00
38	Dibenzo(a,h)anthracene	0.821	0.730	11.1	126	0.00
39	Benzo(g,h,i)perylene	0.827	0.789	4.6	137	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE082018\
Data File : BE097337.D
Acq On : 20 Aug 2018 18:11
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE082018

Quant Time: Aug 20 18:51:11 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE082018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 20 18:32:10 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			