

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE082018\
 Data File : BE097344.D
 Acq On : 20 Aug 2018 22:34
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4EC

Quant Time: Aug 21 05:29:12 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE082018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 20 18:32:10 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	116	0.00
2	1,4-Dioxane	0.400	0.398	0.5	128	0.00
3	n-Nitrosodimethylamine	0.400	0.468	-17.0	158	0.00
4 S	2-Fluorophenol	0.400	0.508	-27.0#	171	0.00
5 S	Phenol-d6	0.400	0.519	-29.8#	171	0.00
6	bis(2-Chloroethyl)ether	0.400	0.443	-10.7	141	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	128	0.01
8 S	Nitrobenzene-d5	0.400	0.567	-41.7#	201	0.00
9	Naphthalene	0.400	0.406	-1.5	139	0.00
10	Hexachlorobutadiene	0.400	0.382	4.5	129	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.439	-9.7	154	0.00
12	2-Methylnaphthalene	0.400	0.435	-8.7	153	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	149	0.00
14 S	2,4,6-Tribromophenol	0.400	0.511	-27.7#	276	0.00
15 S	2-Fluorobiphenyl	0.400	0.370	7.5	146	0.00
16	Acenaphthylene	0.400	0.408	-2.0	177	0.00
17	Acenaphthene	0.400	0.403	-0.8	163	0.00
18	Fluorene	0.400	0.394	1.5	161	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	139	0.00
20	4-Bromophenyl-phenylether	0.400	0.426	-6.5	168	0.00
21	Hexachlorobenzene	0.400	0.382	4.5	139	0.00
22	Pentachlorophenol	0.400	0.643	-60.7#	283	-0.02
23	Phenanthrene	0.400	0.401	-0.3	150	-0.01
24	Anthracene	0.400	0.447	-11.7	172	-0.01
25 SURR	Fluoranthene-d10	0.400	0.436	-9.0	174	0.00
26	Fluoranthene	0.400	0.441	-10.2	165	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	163	-0.01
28	Pyrene	0.400	0.436	-9.0	189	0.00
29 S	Terphenyl-d14	0.400	0.410	-2.5	181	0.00
30	Benzo(a)anthracene	0.400	0.432	-8.0	201	-0.01
31	Chrysene	0.400	0.403	-0.8	174	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.636	-59.0#	272	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.493	-23.2	224	0.00
34 I	Perylene-d12	0.400	0.400	0.0	198	0.00
35	Benzo(b)fluoranthene	0.400	0.363	9.3	200	0.00
36	Benzo(k)fluoranthene	0.400	0.336	16.0	181	0.00
37 C	Benzo(a)pyrene	0.400	0.412	-3.0	226	0.00
38	Dibenzo(a,h)anthracene	0.400	0.373	6.8	191	0.00
39	Benzo(g,h,i)perylene	0.400	0.386	3.5	200	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE082018\
Data File : BE097344.D
Acq On : 20 Aug 2018 22:34
Operator : SJ/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4EC

Quant Time: Aug 21 05:29:12 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE082018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 20 18:32:10 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				