

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE082018\
 Data File : BE097344.D
 Acq On : 20 Aug 2018 22:34
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4EC

Quant Time: Aug 21 05:29:12 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE082018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 20 18:32:10 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.634	0.595	6.2	128	0.00
3	n-Nitrosodimethylamine	0.553	0.648	-17.2	158#	0.00
4 S	2-Fluorophenol	0.917	1.164	-26.9#	171#	0.00
5 S	Phenol-d6	1.359	1.763	-29.7#	171#	0.00
6	bis(2-Chloroethyl)ether	1.353	1.498	-10.7	141	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	128	0.01
8 S	Nitrobenzene-d5	0.171	0.243	-42.1#	201#	0.00
9	Naphthalene	3.666	3.722	-1.5	139	0.00
10	Hexachlorobutadiene	0.161	0.154	4.3	129	0.00
11 SURR	2-Methylnaphthalene-d10	0.541	0.593	-9.6	154#	0.00
12	2-Methylnaphthalene	0.552	0.601	-8.9	153#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	149	0.00
14 S	2,4,6-Tribromophenol	0.066	0.097	-47.0#	276#	0.00
15 S	2-Fluorobiphenyl	1.640	1.515	7.6	146	0.00
16	Acenaphthylene	1.522	1.554	-2.1	177#	0.00
17	Acenaphthene	1.076	1.083	-0.7	163#	0.00
18	Fluorene	1.314	1.295	1.4	161#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	139	0.00
20	4-Bromophenyl-phenylether	0.181	0.192	-6.1	168#	0.00
21	Hexachlorobenzene	0.242	0.231	4.5	139	0.00
22	Pentachlorophenol	0.024	0.030	-25.0	283#	-0.02
23	Phenanthrene	0.976	0.978	-0.2	150#	-0.01
24	Anthracene	0.722	0.808	-11.9	172#	-0.01
25 SURR	Fluoranthene-d10	3.569	3.889	-9.0	174#	0.00
26	Fluoranthene	0.940	1.035	-10.1	165#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	163#	-0.01
28	Pyrene	1.074	1.170	-8.9	189#	0.00
29 S	Terphenyl-d14	0.801	0.821	-2.5	181#	0.00
30	Benzo(a)anthracene	0.860	0.929	-8.0	201#	-0.01
31	Chrysene	1.127	1.135	-0.7	174#	-0.01
32	Bis(2-ethylhexyl)phthalate	0.524	0.756	-44.3#	272#	0.00
33	Indeno(1,2,3-cd)pyrene	0.716	0.882	-23.2	224#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	198#	0.00
35	Benzo(b)fluoranthene	1.055	0.958	9.2	200#	0.00
36	Benzo(k)fluoranthene	1.480	1.245	15.9	181#	0.00
37 C	Benzo(a)pyrene	0.960	0.990	-3.1	226#	0.00
38	Dibenzo(a,h)anthracene	0.821	0.765	6.8	191#	0.00
39	Benzo(g,h,i)perylene	0.827	0.798	3.5	200#	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE082018\
Data File : BE097344.D
Acq On : 20 Aug 2018 22:34
Operator : SJ/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4EC

Quant Time: Aug 21 05:29:12 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE082018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 20 18:32:10 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			