

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE082318\
 Data File : BE097354.D
 Acq On : 23 Aug 2018 10:35
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 23 12:31:14 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE082218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 22 18:07:26 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.711	0.759	-6.8	102	0.00
3	n-Nitrosodimethylamine	0.664	0.649	2.3	96	0.00
4 S	2-Fluorophenol	1.183	1.176	0.6	103	0.00
5 S	Phenol-d6	1.581	1.482	6.3	97	0.00
6	bis(2-Chloroethyl)ether	1.455	1.420	2.4	97	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
8 S	Nitrobenzene-d5	0.287	0.291	-1.4	104	0.00
9	Naphthalene	3.641	3.812	-4.7	102	0.00
10	Hexachlorobutadiene	0.163	0.172	-5.5	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.558	0.571	-2.3	102	0.00
12	2-Methylnaphthalene	0.574	0.593	-3.3	102	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	90	-0.01
14 S	2,4,6-Tribromophenol	0.135	0.125	7.4	100	0.00
15 S	2-Fluorobiphenyl	1.468	1.599	-8.9	103	0.00
16	Acenaphthylene	1.591	1.628	-2.3	103	0.00
17	Acenaphthene	1.061	1.117	-5.3	103	0.00
18	Fluorene	1.342	1.361	-1.4	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
20	4-Bromophenyl-phenylether	0.182	0.194	-6.6	104	0.00
21	Hexachlorobenzene	0.201	0.221	-10.0	104	0.00
22	Pentachlorophenol	0.055	0.042	23.6	88	0.00
23	Phenanthrene	0.952	1.006	-5.7	100	0.00
24	Anthracene	0.810	0.808	0.2	98	0.00
25 SURR	Fluoranthene-d10	4.590	4.821	-5.0	101	0.00
26	Fluoranthene	1.193	1.254	-5.1	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
28	Pvrene	0.952	0.993	-4.3	99	0.00
29 S	Terphenyl-d14	0.738	0.749	-1.5	96	0.00
30	Benzo(a)anthracene	1.012	1.012	0.0	97	0.00
31	Chrysene	1.089	1.147	-5.3	99	0.00
32	Bis(2-ethylhexyl)phthalate	1.361	1.074	21.1	91	-0.01
33	Indeno(1,2,3-cd)pyrene	1.546	1.207	21.9	86	0.00
34 I	Perylene-d12	1.000	1.000	0.0	88	0.00
35	Benzo(b)fluoranthene	1.144	1.025	10.4	93	0.00
36	Benzo(k)fluoranthene	1.264	1.215	3.9	98	0.00
37 C	Benzo(a)pyrene	1.032	0.973	5.7	95	0.00
38	Dibenzo(a,h)anthracene	1.356	1.074	20.8	86	0.00
39	Benzo(a,h,i)perylene	1.400	1.078	23.0	84	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE082318\
Data File : BE097354.D
Acq On : 23 Aug 2018 10:35
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 23 12:31:14 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE082218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 22 18:07:26 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			