

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE082918\
 Data File : BE097387.D
 Acq On : 29 Aug 2018 13:37
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SICV19

Quant Time: Aug 29 14:21:11 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE082918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 29 13:18:39 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
3	Naphthalene	0.975	0.973	0.2	99	0.00
4 SURR	2-Methylnaphthalene-d10	0.622	0.627	-0.8	99	0.00
5	2-Methylnaphthalene	0.616	0.618	-0.3	99	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
7	Acenaphthylene	1.457	1.436	1.4	98	0.00
8 C	Acenaphthene	1.253	1.243	0.8	98	0.00
9	Fluorene	1.435	1.447	-0.8	100	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
11	Pentachlorophenol	0.039	0.033	15.4	103	0.00
12	Phenanthrene	1.030	1.031	-0.1	100	0.00
13	Anthracene	0.866	0.868	-0.2	102	0.00
14 SURR	Fluoranthene-d10	1.319	1.256	4.8	98	0.00
15 C	Fluoranthene	1.323	1.266	4.3	98	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
17	Pyrene	1.076	1.082	-0.6	101	0.00
18	Benzo(a)anthracene	1.028	1.023	0.5	101	0.00
19	Chrysene	1.184	1.182	0.2	100	0.00
20 I	Perylene-d12	1.000	1.000	0.0	96	0.00
21	Benzo(b)fluoranthene	1.071	1.039	3.0	99	0.00
22	Benzo(k)fluoranthene	1.195	1.211	-1.3	97	0.00
23 C	Benzo(a)pyrene	0.991	0.958	3.3	97	0.00
24	Indeno(1,2,3-cd)pyrene	1.284	1.270	1.1	96	0.00
25	Dibenzo(a,h)anthracene	1.088	1.078	0.9	95	0.00
26	Benzo(g,h,i)perylene	1.095	1.081	1.3	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0