

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE082918\
 Data File : BE097399.D
 Acq On : 29 Aug 2018 21:08
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.421

Quant Time: Aug 30 02:13:02 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE082918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 30 02:12:09 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
3	Naphthalene	0.975	0.981	-0.6	102	0.00
4 SURR	2-Methylnaphthalene-d10	0.622	0.664	-6.8	106	0.00
5	2-Methylnaphthalene	0.616	0.652	-5.8	106	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
7	Acenaphthylene	1.457	1.562	-7.2	118	0.00
8 C	Acenaphthene	1.253	1.263	-0.8	110	0.00
9	Fluorene	1.435	1.513	-5.4	115	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
11	Pentachlorophenol	0.039	0.050	-28.2#	170#	0.00
12	Phenanthrene	1.030	1.034	-0.4	110	0.00
13	Anthracene	0.866	0.932	-7.6	120	0.00
14 SURR	Fluoranthene-d10	1.319	1.389	-5.3	119	0.00
15 C	Fluoranthene	1.323	1.367	-3.3	116	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	116	0.00
17	Pyrene	1.076	1.072	0.4	117	0.00
18	Benzo(a)anthracene	1.028	1.166	-13.4	135	0.00
19	Chrysene	1.184	1.199	-1.3	118	0.00
20 I	Perylene-d12	1.000	1.000	0.0	121	0.00
21	Benzo(b)fluoranthene	1.071	1.164	-8.7	139	0.00
22	Benzo(k)fluoranthene	1.195	1.150	3.8	116	0.00
23 C	Benzo(a)pyrene	0.991	1.105	-11.5	141	0.00
24	Indeno(1,2,3-cd)pyrene	1.284	1.342	-4.5	127	0.00
25	Dibenzo(a,h)anthracene	1.088	1.116	-2.6	124	0.00
26	Benzo(g,h,i)perylene	1.095	1.125	-2.7	126	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0