

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE090718\
 Data File : BE097428.D
 Acq On : 7 Sep 2018 19:03
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.427

Quant Time: Sep 08 02:46:03 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE082918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 07 23:49:56 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
3	Naphthalene	0.975	0.978	-0.3	109	0.00
4 SURR	2-Methylnaphthalene-d10	0.622	0.685	-10.1	118	0.00
5	2-Methylnaphthalene	0.616	0.680	-10.4	119	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	125	0.00
7	Acenaphthylene	1.457	1.618	-11.1	141	0.00
8 C	Acenaphthene	1.253	1.267	-1.1	127	0.00
9	Fluorene	1.435	1.504	-4.8	131	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
11	Pentachlorophenol	0.039	0.057	-46.2#	214#	0.00
12	Phenanthrene	1.030	1.041	-1.1	121	0.00
13	Anthracene	0.866	0.965	-11.4	136	0.00
14 SURR	Fluoranthene-d10	1.319	1.325	-0.5	124	0.00
15 C	Fluoranthene	1.323	1.287	2.7	119	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
17	Pyrene	1.076	1.115	-3.6	118	0.00
18	Benzo(a)anthracene	1.028	1.159	-12.7	130	0.00
19	Chrysene	1.184	1.168	1.4	112	0.00
20 I	Perylene-d12	1.000	1.000	0.0	116	0.00
21	Benzo(b)fluoranthene	1.071	1.153	-7.7	132	0.00
22	Benzo(k)fluoranthene	1.195	1.135	5.0	110	0.00
23 C	Benzo(a)pyrene	0.991	1.103	-11.3	135	0.00
24	Indeno(1,2,3-cd)pyrene	1.284	1.382	-7.6	126	0.00
25	Dibenzo(a,h)anthracene	1.088	1.154	-6.1	124	0.00
26	Benzo(g,h,i)perylene	1.095	1.155	-5.5	125	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0