

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE091018\
 Data File : BE097440.D
 Acq On : 10 Sep 2018 8:51
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 10 09:37:00 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE082218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 22 18:07:26 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	98	0.00
2	1,4-Dioxane	0.400	0.362	9.5	90	0.00
3	n-Nitrosodimethylamine	0.400	0.349	12.8	90	0.00
4 S	2-Fluorophenol	0.400	0.426	-6.5	116	-0.01
5 S	Phenol-d6	0.400	0.455	-13.7	123	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.378	5.5	99	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	109	0.00
8 S	Nitrobenzene-d5	0.400	0.443	-10.7	135	-0.01
9	Naphthalene	0.400	0.415	-3.7	120	-0.01
10	Hexachlorobutadiene	0.400	0.418	-4.5	120	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.458	-14.5	135	0.00
12	2-Methylnaphthalene	0.400	0.453	-13.2	133	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	132	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.523	-30.8#	209	0.00
15 S	2-Fluorobiphenyl	0.400	0.392	2.0	136	0.00
16	Acenaphthylene	0.400	0.437	-9.2	161	-0.01
17	Acenaphthene	0.400	0.432	-8.0	155	0.00
18	Fluorene	0.400	0.431	-7.7	155	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	125	-0.01
20	4-Bromophenyl-phenylether	0.400	0.438	-9.5	153	0.00
21	Hexachlorobenzene	0.400	0.445	-11.2	151	-0.01
22	Pentachlorophenol	0.400	0.493	-23.2	170	-0.01
23	Phenanthrene	0.400	0.427	-6.7	145	0.00
24	Anthracene	0.400	0.441	-10.2	155	0.00
25 SURR	Fluoranthene-d10	0.400	0.393	1.8	135	0.00
26	Fluoranthene	0.400	0.390	2.5	135	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	107	0.00
28	Pvrene	0.400	0.472	-18.0	136	0.00
29 S	Terphenyl-d14	0.400	0.440	-10.0	127	0.00
30	Benzo(a)anthracene	0.400	0.437	-9.2	129	0.00
31	Chrysene	0.400	0.411	-2.7	117	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.621	-55.2#	200	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.377	5.8	113	0.00
34 I	Perylene-d12	0.400	0.400	0.0	112	0.00
35	Benzo(b)fluoranthene	0.400	0.389	2.8	121	0.00
36	Benzo(k)fluoranthene	0.400	0.387	3.3	119	0.00
37 C	Benzo(a)pyrene	0.400	0.398	0.5	129	0.00
38	Dibenzo(a,h)anthracene	0.400	0.355	11.3	112	0.00
39	Benzo(a,h,i)perylene	0.400	0.340	15.0	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE091018\
Data File : BE097440.D
Acq On : 10 Sep 2018 8:51
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 10 09:37:00 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE082218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 22 18:07:26 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				