

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE091018\
 Data File : BE097440.D
 Acq On : 10 Sep 2018 8:51
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 10 09:37:00 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE082218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 22 18:07:26 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.711	0.642	9.7	90	0.00
3	n-Nitrosodimethylamine	0.664	0.579	12.8	90	0.00
4 S	2-Fluorophenol	1.183	1.258	-6.3	116	-0.01
5 S	Phenol-d6	1.581	1.797	-13.7	123	-0.01
6	bis(2-Chloroethyl)ether	1.455	1.374	5.6	99	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
8 S	Nitrobenzene-d5	0.287	0.318	-10.8	135	-0.01
9	Naphthalene	3.641	3.781	-3.8	120	-0.01
10	Hexachlorobutadiene	0.163	0.170	-4.3	120	0.00
11 SURR	2-Methylnaphthalene-d10	0.558	0.639	-14.5	135	0.00
12	2-Methylnaphthalene	0.574	0.650	-13.2	133	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	132	-0.01
14 S	2,4,6-Tribromophenol	0.135	0.177	-31.1#	209#	0.00
15 S	2-Fluorobiphenyl	1.468	1.440	1.9	136	0.00
16	Acenaphthylene	1.591	1.738	-9.2	161#	-0.01
17	Acenaphthene	1.061	1.146	-8.0	155#	0.00
18	Fluorene	1.342	1.448	-7.9	155#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	125	-0.01
20	4-Bromophenyl-phenylether	0.182	0.199	-9.3	153#	0.00
21	Hexachlorobenzene	0.201	0.224	-11.4	151#	-0.01
22	Pentachlorophenol	0.055	0.057	-3.6	170#	-0.01
23	Phenanthrene	0.952	1.015	-6.6	145	0.00
24	Anthracene	0.810	0.893	-10.2	155#	0.00
25 SURR	Fluoranthene-d10	4.590	4.508	1.8	135	0.00
26	Fluoranthene	1.193	1.164	2.4	135	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
28	Pvrene	0.952	1.122	-17.9	136	0.00
29 S	Terphenyl-d14	0.738	0.813	-10.2	127	0.00
30	Benzo(a)anthracene	1.012	1.105	-9.2	129	0.00
31	Chrysene	1.089	1.120	-2.8	117	0.00
32	Bis(2-ethylhexyl)phthalate	1.361	1.949	-43.2#	200#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.546	1.303	15.7	113	0.00
34 I	Perylene-d12	1.000	1.000	0.0	112	0.00
35	Benzo(b)fluoranthene	1.144	1.041	9.0	121	0.00
36	Benzo(k)fluoranthene	1.264	1.146	9.3	119	0.00
37 C	Benzo(a)pyrene	1.032	1.026	0.6	129	0.00
38	Dibenzo(a,h)anthracene	1.356	1.091	19.5	112	0.00
39	Benzo(a,h,i)perylene	1.400	1.085	22.5	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE091018\
Data File : BE097440.D
Acq On : 10 Sep 2018 8:51
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 10 09:37:00 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE082218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 22 18:07:26 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		