

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE091018\
 Data File : BE097445.D
 Acq On : 10 Sep 2018 13:37
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 10 17:07:40 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE082218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 22 18:07:26 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.711	0.661	7.0	92	0.00
3	n-Nitrosodimethylamine	0.664	0.628	5.4	97	0.00
4 S	2-Fluorophenol	1.183	1.214	-2.6	111	-0.01
5 S	Phenol-d6	1.581	1.706	-7.9	116	-0.01
6	bis(2-Chloroethyl)ether	1.455	1.382	5.0	98	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	110	0.00
8 S	Nitrobenzene-d5	0.287	0.307	-7.0	132	0.00
9	Naphthalene	3.641	3.748	-2.9	120	-0.01
10	Hexachlorobutadiene	0.163	0.164	-0.6	117	0.00
11 SURR	2-Methylnaphthalene-d10	0.558	0.637	-14.2	136	0.00
12	2-Methylnaphthalene	0.574	0.657	-14.5	135	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	139	-0.01
14 S	2,4,6-Tribromophenol	0.135	0.156	-15.6	194#	0.00
15 S	2-Fluorobiphenyl	1.468	1.415	3.6	141	0.00
16	Acenaphthylene	1.591	1.669	-4.9	163#	-0.01
17	Acenaphthene	1.061	1.092	-2.9	156#	-0.01
18	Fluorene	1.342	1.441	-7.4	162#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	130	-0.01
20	4-Bromophenyl-phenylether	0.182	0.198	-8.8	158#	0.00
21	Hexachlorobenzene	0.201	0.221	-10.0	155#	-0.01
22	Pentachlorophenol	0.055	0.051	7.3	160#	-0.01
23	Phenanthrene	0.952	1.000	-5.0	149	0.00
24	Anthracene	0.810	0.885	-9.3	160#	0.00
25 SURR	Fluoranthene-d10	4.590	4.393	4.3	137	0.00
26	Fluoranthene	1.193	1.126	5.6	136	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
28	Pvrene	0.952	1.119	-17.5	135	0.00
29 S	Terphenyl-d14	0.738	0.807	-9.3	125	0.00
30	Benzo(a)anthracene	1.012	1.065	-5.2	123	0.00
31	Chrysene	1.089	1.128	-3.6	118	0.00
32	Bis(2-ethylhexyl)phthalate	1.361	1.509	-10.9	154#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.546	1.287	16.8	111	0.00
34 I	Perylene-d12	1.000	1.000	0.0	108	0.00
35	Benzo(b)fluoranthene	1.144	1.102	3.7	123	0.00
36	Benzo(k)fluoranthene	1.264	1.191	5.8	118	0.00
37 C	Benzo(a)pyrene	1.032	1.040	-0.8	125	0.00
38	Dibenzo(a,h)anthracene	1.356	1.114	17.8	109	0.00
39	Benzo(a,h,i)perylene	1.400	1.088	22.3	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE091018\
Data File : BE097445.D
Acq On : 10 Sep 2018 13:37
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 10 17:07:40 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE082218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 22 18:07:26 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			