

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE091718\
 Data File : BE097503.D
 Acq On : 17 Sep 2018 19:37
 Operator : SJ/JU
 Sample : J4892-02
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C-MW-06-091118

Manual Integrations
 APPROVED

Sohil
 9/18/2018 5:56:03 PM

Quant Time: Sep 18 15:00:26 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 17 18:57:11 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.38	152	951	0.40	ng	0.00
7) Naphthalene-d8	10.13	136	4406	0.40	ng	0.00
13) Acenaphthene-d10	14.00	164	2576	0.40	ng	0.00
19) Phenanthrene-d10	16.77	188	7283	0.40	ng	0.00
27) Chrysene-d12	20.97	240	9433	0.40	ng	0.00
34) Perylene-d12	23.21	264	9461	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.05	112	683	0.21	ng	0.00
5) Phenol-d6	6.60	99	506	0.12	ng	0.00
8) Nitrobenzene-d5	8.53	82	1574	0.44	ng	0.00
11) 2-Methylnaphthalene-d10	11.72	152	2998	0.48	ng	0.00
14) 2,4,6-Tribromophenol	15.52	330	610	0.45	ng	-0.01
15) 2-Fluorobiphenyl	12.62	172	5368	0.60	ng	0.00
25) Fluoranthene-d10	18.80	212	46136	0.53	ng	0.00
29) Terphenyl-d14	19.42	244	11718	0.65	ng	0.00
Target Compounds						
22) Pentachlorophenol	16.43	266	149m	0.340	ng	Qvalue
26) Fluoranthene	18.83	202	579	0.026	ng	# 93
28) Pyrene	19.20	202	540	0.022	ng	96
32) Bis(2-ethylhexyl)phthalate	20.91	149	11104	0.211	ng	# 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE091718\
Data File : BE097503.D
Acq On : 17 Sep 2018 19:37
Operator : SJ/JU
Sample : J4892-02
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C-MW-06-091118

Manual Integrations
APPROVED

Sohil
9/18/2018 5:56:03 PM

Quant Time: Sep 18 15:00:26 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 17 18:57:11 2018
Response via : Initial Calibration

