

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE091718\
 Data File : BE097508.D
 Acq On : 17 Sep 2018 22:42
 Operator : SJ/JU
 Sample : J4892-10
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PR-MW-03-091218

Quant Time: Sep 18 15:12:39 2018

Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091718.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Mon Sep 17 18:57:11 2018

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.37	152	996	0.40	ng	0.00
7) Naphthalene-d8	10.12	136	5081	0.40	ng	-0.01
13) Acenaphthene-d10	14.00	164	5216	0.40	ng	0.00
19) Phenanthrene-d10	16.75	188	8085	0.40	ng	-0.01
27) Chrysene-d12	20.97	240	9164	0.40	ng	0.00
34) Perylene-d12	23.21	264	9221	0.40	ng	0.00

System Monitoring Compounds

4) 2-Fluorophenol	5.05	112	668	0.20	ng	0.00
5) Phenol-d6	6.60	99	526	0.12	ng	0.00
8) Nitrobenzene-d5	8.51	82	1640	0.40	ng	-0.02
11) 2-Methylnaphthalene-d10	11.71	152	3276	0.45	ng	0.00
14) 2,4,6-Tribromophenol	15.51	330	764	0.28	ng	-0.02
15) 2-Fluorobiphenyl	12.61	172	6028	0.33	ng	-0.01
25) Fluoranthene-d10	18.80	212	43857	0.45	ng	0.00
29) Terphenyl-d14	19.41	244	10635	0.61	ng	0.00

Target Compounds

					Qvalue	
9) Naphthalene	10.17	128	53268	1.155	ng	# 93
12) 2-Methylnaphthalene	11.78	142	151837	20.580	ng	99
17) Acenaphthene	14.06	154	1839	0.131	ng	# 48
18) Fluorene	15.05	166	6554	0.367	ng	# 78
23) Phenanthrene	16.79	178	2444	0.129	ng	# 96
32) Bis(2-ethylhexyl)phthalate	20.91	149	6340	0.124	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE091718\
Data File : BE097508.D
Acq On : 17 Sep 2018 22:42
Operator : SJ/JU
Sample : J4892-10
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PR-MW-03-091218

Quant Time: Sep 18 15:12:39 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 17 18:57:11 2018
Response via : Initial Calibration

