

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE091718\
 Data File : BE097526.D
 Acq On : 18 Sep 2018 10:36
 Operator : SJ/JU
 Sample : J4913-05
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 BPSI-TT-MW30312-20180912

Manual Integrations
 APPROVED

Sohil
 9/18/2018 5:56:35 PM

Quant Time: Sep 18 11:46:08 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 17 18:57:11 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.38	152	974	0.40	ng	0.00
7) Naphthalene-d8	10.13	136	4402	0.40	ng	0.00
13) Acenaphthene-d10	14.00	164	2559	0.40	ng	0.00
19) Phenanthrene-d10	16.75	188	7473	0.40	ng	0.00
27) Chrysene-d12	20.97	240	8902	0.40	ng	0.00
34) Perylene-d12	23.21	264	8878	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.05	112	589	0.18	ng	0.00
5) Phenol-d6	6.61	99	496m	0.12	ng	0.00
8) Nitrobenzene-d5	8.53	82	1497	0.42	ng	0.00
11) 2-Methylnaphthalene-d10	11.72	152	2674	0.43	ng	0.00
14) 2,4,6-Tribromophenol	15.52	330	520	0.38	ng	0.00
15) 2-Fluorobiphenyl	12.62	172	5114	0.57	ng	0.00
25) Fluoranthene-d10	18.80	212	39586	0.44	ng	0.00
29) Terphenyl-d14	19.41	244	8742	0.51	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.08	88	1054	0.626	ng	Qvalue 95
32) Bis(2-ethylhexyl)phthalate	20.91	149	3969	0.080	ng	# 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE091718\
Data File : BE097526.D
Acq On : 18 Sep 2018 10:36
Operator : SJ/JU
Sample : J4913-05
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
BPSI-TT-MW30312-20180912

Manual Integrations
APPROVED

Sohil
9/18/2018 5:56:35 PM

Quant Time: Sep 18 11:46:08 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 17 18:57:11 2018
Response via : Initial Calibration

