

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE091718\
 Data File : BE097529.D
 Acq On : 18 Sep 2018 12:29
 Operator : SJ/JU
 Sample : J4913-08
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 BP-DUP01-20180912

Manual Integrations
 APPROVED

Sohil
 9/18/2018 5:56:37 PM

Quant Time: Sep 18 13:16:48 2018

Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091718.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Mon Sep 17 18:57:11 2018

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.38	152	991	0.40	ng	0.00
7) Naphthalene-d8	10.13	136	4609	0.40	ng	0.00
13) Acenaphthene-d10	14.00	164	2729	0.40	ng	0.00
19) Phenanthrene-d10	16.75	188	7416	0.40	ng	-0.01
27) Chrysene-d12	20.97	240	8615	0.40	ng	0.00
34) Perylene-d12	23.21	264	8702	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.05	112	519	0.15	ng	0.00
5) Phenol-d6	6.62	99	361m	0.08	ng	0.01
8) Nitrobenzene-d5	8.54	82	1392	0.38	ng	0.01
11) 2-Methylnaphthalene-d10	11.72	152	2631	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.53	330	538	0.37	ng	0.00
15) 2-Fluorobiphenyl	12.62	172	4860	0.51	ng	0.00
25) Fluoranthene-d10	18.80	212	35611	0.40	ng	0.00
29) Terphenyl-d14	19.41	244	7955	0.48	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.08	88	870	0.508	ng	94
32) Bis(2-ethylhexyl)phthalate	20.91	149	1954	0.041	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE091718\
Data File : BE097529.D
Acq On : 18 Sep 2018 12:29
Operator : SJ/JU
Sample : J4913-08
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
BP-DUP01-20180912

Manual Integrations
APPROVED

Sohil
9/18/2018 5:56:37 PM

Quant Time: Sep 18 13:16:48 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 17 18:57:11 2018
Response via : Initial Calibration

