

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091718\
 Data File : BE097495.D
 Acq On : 17 Sep 2018 14:01
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE091718

Quant Time: Sep 17 18:57:25 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 17 18:57:11 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	108	0.00
2	1,4-Dioxane	0.400	0.393	1.8	111	0.00
3	n-Nitrosodimethylamine	0.400	0.425	-6.2	140	0.00
4 S	2-Fluorophenol	0.400	0.389	2.8	119	0.00
5 S	Phenol-d6	0.400	0.385	3.8	130	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.400	0.0	116	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	111	0.00
8 S	Nitrobenzene-d5	0.400	0.373	6.8	125	-0.01
9	Naphthalene	0.400	0.411	-2.7	123	0.00
10	Hexachlorobutadiene	0.400	0.405	-1.3	121	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.405	-1.3	122	0.00
12	2-Methylnaphthalene	0.400	0.410	-2.5	125	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	112	0.00
14 S	2,4,6-Tribromophenol	0.400	0.364	9.0	126	0.00
15 S	2-Fluorobiphenyl	0.400	0.411	-2.7	124	0.00
16	Acenaphthylene	0.400	0.377	5.8	118	0.00
17	Acenaphthene	0.400	0.409	-2.2	123	0.00
18	Fluorene	0.400	0.406	-1.5	122	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	111	0.00
20	4-Bromophenyl-phenylether	0.400	0.381	4.8	119	0.00
21	Hexachlorobenzene	0.400	0.398	0.5	120	0.00
22	Pentachlorophenol	0.400	0.378	5.5	122	-0.02
23	Phenanthrene	0.400	0.411	-2.7	123	0.00
24	Anthracene	0.400	0.395	1.3	121	0.00
25 SURR	Fluoranthene-d10	0.400	0.396	1.0	121	0.00
26	Fluoranthene	0.400	0.397	0.8	121	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	110	0.00
28	Pyrene	0.400	0.414	-3.5	124	0.00
29 S	Terphenyl-d14	0.400	0.415	-3.7	128	0.00
30	Benzo(a)anthracene	0.400	0.388	3.0	121	0.00
31	Chrysene	0.400	0.418	-4.5	120	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.388	3.0	113	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.412	-3.0	122	0.00
34 I	Perylene-d12	0.400	0.400	0.0	111	0.00
35	Benzo(b)fluoranthene	0.400	0.407	-1.7	125	0.00
36	Benzo(k)fluoranthene	0.400	0.406	-1.5	120	0.00
37 C	Benzo(a)pyrene	0.400	0.403	-0.8	121	0.00
38	Dibenzo(a,h)anthracene	0.400	0.412	-3.0	125	0.00
39	Benzo(g,h,i)perylene	0.400	0.399	0.3	118	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091718\
Data File : BE097495.D
Acq On : 17 Sep 2018 14:01
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE091718

Quant Time: Sep 17 18:57:25 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 17 18:57:11 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				