

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091718\
 Data File : BE097495.D
 Acq On : 17 Sep 2018 14:01
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE091718

Quant Time: Sep 17 18:57:25 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 17 18:57:11 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.692	0.679	1.9	111	0.00
3	n-Nitrosodimethylamine	0.578	0.614	-6.2	140	0.00
4 S	2-Fluorophenol	1.355	1.319	2.7	119	0.00
5 S	Phenol-d6	1.719	1.656	3.7	130	-0.01
6	bis(2-Chloroethyl)ether	1.374	1.374	0.0	116	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
8 S	Nitrobenzene-d5	0.322	0.301	6.5	125	-0.01
9	Naphthalene	3.631	3.735	-2.9	123	0.00
10	Hexachlorobutadiene	0.172	0.174	-1.2	121	0.00
11 SURR	2-Methylnaphthalene-d10	0.572	0.579	-1.2	122	0.00
12	2-Methylnaphthalene	0.581	0.595	-2.4	125	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
14 S	2,4,6-Tribromophenol	0.211	0.192	9.0	126	0.00
15 S	2-Fluorobiphenyl	1.397	1.436	-2.8	124	0.00
16	Acenaphthylene	1.740	1.639	5.8	118	0.00
17	Acenaphthene	1.074	1.099	-2.3	123	0.00
18	Fluorene	1.369	1.389	-1.5	122	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
20	4-Bromophenyl-phenylether	0.193	0.184	4.7	119	0.00
21	Hexachlorobenzene	0.205	0.204	0.5	120	0.00
22	Pentachlorophenol	0.036	0.023	36.1#	122	-0.02
23	Phenanthrene	0.941	0.967	-2.8	123	0.00
24	Anthracene	0.869	0.859	1.2	121	0.00
25 SURR	Fluoranthene-d10	4.806	4.754	1.1	121	0.00
26	Fluoranthene	1.227	1.217	0.8	121	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
28	Pyrene	1.029	1.066	-3.6	124	0.00
29 S	Terphenyl-d14	0.765	0.794	-3.8	128	0.00
30	Benzo(a)anthracene	1.052	1.021	2.9	121	0.00
31	Chrysene	1.079	1.128	-4.5	120	0.00
32	Bis(2-ethylhexyl)phthalate	2.235	2.168	3.0	113	0.00
33	Indeno(1,2,3-cd)pyrene	1.286	1.326	-3.1	122	0.00
34 I	Perylene-d12	1.000	1.000	0.0	111	0.00
35	Benzo(b)fluoranthene	1.023	1.040	-1.7	125	0.00
36	Benzo(k)fluoranthene	1.115	1.132	-1.5	120	0.00
37 C	Benzo(a)pyrene	1.016	1.025	-0.9	121	0.00
38	Dibenzo(a,h)anthracene	1.060	1.092	-3.0	125	0.00
39	Benzo(g,h,i)perylene	1.082	1.079	0.3	118	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091718\
Data File : BE097495.D
Acq On : 17 Sep 2018 14:01
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE091718

Quant Time: Sep 17 18:57:25 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 17 18:57:11 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		