

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091718\
 Data File : BE097514.D
 Acq On : 18 Sep 2018 2:29
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 18 07:11:39 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 17 18:57:11 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.692	0.681	1.6	98	0.00
3	n-Nitrosodimethylamine	0.578	0.598	-3.5	120	0.00
4 S	2-Fluorophenol	1.355	1.326	2.1	104	0.00
5 S	Phenol-d6	1.719	1.789	-4.1	123	-0.01
6	bis(2-Chloroethyl)ether	1.374	1.370	0.3	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
8 S	Nitrobenzene-d5	0.322	0.316	1.9	117	-0.01
9	Naphthalene	3.631	3.736	-2.9	109	0.00
10	Hexachlorobutadiene	0.172	0.169	1.7	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.572	0.587	-2.6	110	0.00
12	2-Methylnaphthalene	0.581	0.603	-3.8	113	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.211	0.182	13.7	106	-0.01
15 S	2-Fluorobiphenyl	1.397	1.460	-4.5	112	0.00
16	Acenaphthylene	1.740	1.707	1.9	109	0.00
17	Acenaphthene	1.074	1.124	-4.7	112	0.00
18	Fluorene	1.369	1.472	-7.5	115	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	-0.01
20	4-Bromophenyl-phenylether	0.193	0.192	0.5	111	0.00
21	Hexachlorobenzene	0.205	0.213	-3.9	111	0.00
22	Pentachlorophenol	0.036	0.070	-94.4#	321#	-0.05
23	Phenanthrene	0.941	0.990	-5.2	112	0.00
24	Anthracene	0.869	0.860	1.0	108	0.00
25 SURR	Fluoranthene-d10	4.806	4.676	2.7	106	0.00
26	Fluoranthene	1.227	1.194	2.7	105	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
28	Pyrene	1.029	1.072	-4.2	103	0.00
29 S	Terphenyl-d14	0.765	0.805	-5.2	108	0.00
30	Benzo(a)anthracene	1.052	1.070	-1.7	105	0.00
31	Chrysene	1.079	1.101	-2.0	97	0.00
32	Bis(2-ethylhexyl)phthalate	2.235	2.183	2.3	94	0.00
33	Indeno(1,2,3-cd)pyrene	1.286	1.385	-7.7	106	0.00
34 I	Perylene-d12	1.000	1.000	0.0	94	0.00
35	Benzo(b)fluoranthene	1.023	1.012	1.1	103	0.00
36	Benzo(k)fluoranthene	1.115	1.176	-5.5	106	0.00
37 C	Benzo(a)pyrene	1.016	1.034	-1.8	104	0.00
38	Dibenzo(a,h)anthracene	1.060	1.126	-6.2	109	0.00
39	Benzo(g,h,i)perylene	1.082	1.095	-1.2	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091718\
Data File : BE097514.D
Acq On : 18 Sep 2018 2:29
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 18 07:11:39 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 17 18:57:11 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			