

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091718\
 Data File : BE097533.D
 Acq On : 18 Sep 2018 14:58
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 18 16:21:52 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 17 18:57:11 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.692	0.650	6.1	110	0.00
3	n-Nitrosodimethylamine	0.578	0.537	7.1	127	-0.01
4 S	2-Fluorophenol	1.355	1.238	8.6	115	0.00
5 S	Phenol-d6	1.719	1.688	1.8	137	-0.01
6	bis(2-Chloroethyl)ether	1.374	1.388	-1.0	122	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	117	-0.01
8 S	Nitrobenzene-d5	0.322	0.307	4.7	134	-0.01
9	Naphthalene	3.631	3.776	-4.0	130	0.00
10	Hexachlorobutadiene	0.172	0.170	1.2	123	0.00
11 SURR	2-Methylnaphthalene-d10	0.572	0.587	-2.6	129	0.00
12	2-Methylnaphthalene	0.581	0.608	-4.6	134	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.211	0.177	16.1	124	-0.01
15 S	2-Fluorobiphenyl	1.397	1.449	-3.7	133	0.00
16	Acenaphthylene	1.740	1.676	3.7	129	0.00
17	Acenaphthene	1.074	1.114	-3.7	134	0.00
18	Fluorene	1.369	1.424	-4.0	134	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	118	-0.01
20	4-Bromophenyl-phenylether	0.193	0.192	0.5	133	0.00
21	Hexachlorobenzene	0.205	0.214	-4.4	134	0.00
22	Pentachlorophenol	0.036	0.069	-91.7#	383#	-0.05
23	Phenanthrene	0.941	0.992	-5.4	134	0.00
24	Anthracene	0.869	0.840	3.3	126	0.00
25 SURR	Fluoranthene-d10	4.806	4.663	3.0	126	0.00
26	Fluoranthene	1.227	1.205	1.8	127	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
28	Pyrene	1.029	1.094	-6.3	125	0.00
29 S	Terphenyl-d14	0.765	0.819	-7.1	130	0.00
30	Benzo(a)anthracene	1.052	1.049	0.3	122	0.00
31	Chrysene	1.079	1.149	-6.5	120	0.00
32	Bis(2-ethylhexyl)phthalate	2.235	1.837	17.8	94	0.00
33	Indeno(1,2,3-cd)pyrene	1.286	1.430	-11.2	130	0.00
34 I	Perylene-d12	1.000	1.000	0.0	114	0.00
35	Benzo(b)fluoranthene	1.023	0.997	2.5	122	0.00
36	Benzo(k)fluoranthene	1.115	1.226	-10.0	133	0.00
37 C	Benzo(a)pyrene	1.016	1.031	-1.5	125	0.00
38	Dibenzo(a,h)anthracene	1.060	1.134	-7.0	132	0.00
39	Benzo(g,h,i)perylene	1.082	1.090	-0.7	122	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091718\
Data File : BE097533.D
Acq On : 18 Sep 2018 14:58
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 18 16:21:52 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 17 18:57:11 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		