

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092018\
 Data File : BE097579.D
 Acq On : 20 Sep 2018 18:24
 Operator : SJ/JU
 Sample : J4913-15
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 BPS1-TT-MW306D-20180913

Quant Time: Sep 21 06:40:55 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 20 10:54:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.38	152	686	0.40	ng	0.00
7) Naphthalene-d8	10.15	136	3156	0.40	ng	0.01
13) Acenaphthene-d10	14.01	164	2004	0.40	ng	0.01
19) Phenanthrene-d10	16.76	188	6211	0.40	ng	0.00
27) Chrysene-d12	20.97	240	7900	0.40	ng	0.00
34) Perylene-d12	23.21	264	7838	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.07	112	16	0.01	ng	0.01
5) Phenol-d6	6.62	99	3	0.00	ng	0.01
8) Nitrobenzene-d5	8.56	82	890	0.37	ng	0.02
11) 2-Methylnaphthalene-d10	11.74	152	1981	0.46	ng	0.02
14) 2,4,6-Tribromophenol	15.55	330	51	0.15	ng	0.03
15) 2-Fluorobiphenyl	12.64	172	4149	0.59	ng	0.02
25) Fluoranthene-d10	18.80	212	36533	0.46	ng	0.00
29) Terphenyl-d14	19.42	244	8034	0.56	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.07	88	2030	1.888	ng	# 94
32) Bis(2-ethylhexyl)phthalate	20.91	149	1237	0.047	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092018\
Data File : BE097579.D
Acq On : 20 Sep 2018 18:24
Operator : SJ/JU
Sample : J4913-15
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
BPS1-TT-MW306D-20180913

Quant Time: Sep 21 06:40:55 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 20 10:54:43 2018
Response via : Initial Calibration

