

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092018\
 Data File : BE097593.D
 Acq On : 21 Sep 2018 3:42
 Operator : SJ/JU
 Sample : J4963-06
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SA-PZ122D-20180914

Manual Integrations
 APPROVED

Sohil
 9/21/2018 4:09:35 PM

Quant Time: Sep 21 07:11:01 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 20 10:54:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.38	152	808	0.40	ng	0.00
7) Naphthalene-d8	10.13	136	3652	0.40	ng	0.00
13) Acenaphthene-d10	14.00	164	2221	0.40	ng	0.00
19) Phenanthrene-d10	16.75	188	6922	0.40	ng	0.00
27) Chrysene-d12	20.97	240	9619	0.40	ng	0.00
34) Perylene-d12	23.20	264	9395	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.05	112	464	0.19	ng	0.00
5) Phenol-d6	6.66	99	364m	0.12	ng	0.05
8) Nitrobenzene-d5	8.56	82	1190	0.42	ng	0.03
11) 2-Methylnaphthalene-d10	11.72	152	2372	0.47	ng	0.00
14) 2,4,6-Tribromophenol	15.52	330	617	0.66	ng	0.00
15) 2-Fluorobiphenyl	12.62	172	5388	0.69	ng	0.00
25) Fluoranthene-d10	18.80	212	43298	0.49	ng	0.00
29) Terphenyl-d14	19.41	244	9942	0.56	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.07	88	909	0.689	ng	Qvalue 94
9) Naphthalene	10.18	128	963	0.029	ng	# 78
32) Bis(2-ethylhexyl)phthalate	20.89	149	1050	0.036	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092018\
Data File : BE097593.D
Acq On : 21 Sep 2018 3:42
Operator : SJ/JU
Sample : J4963-06
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SA-PZ122D-20180914

Manual Integrations
APPROVED

Sohil
9/21/2018 4:09:35 PM

Quant Time: Sep 21 07:11:01 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 20 10:54:43 2018
Response via : Initial Calibration

