

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092018\
 Data File : BE097604.D
 Acq On : 21 Sep 2018 10:28
 Operator : SJ/JU
 Sample : J4892-10
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PR-MW-03-091218

Quant Time: Sep 21 13:43:38 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 20 10:54:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.37	152	742	0.40	ng	-0.01
7) Naphthalene-d8	10.12	136	3868	0.40	ng	-0.01
13) Acenaphthene-d10	13.99	164	3867	0.40	ng	-0.01
19) Phenanthrene-d10	16.75	188	7178	0.40	ng	0.00
27) Chrysene-d12	20.97	240	9677	0.40	ng	0.00
34) Perylene-d12	23.20	264	9444	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.05	112	505	0.23	ng	0.00
5) Phenol-d6	6.59	99	356	0.13	ng	-0.01
8) Nitrobenzene-d5	8.52	82	1165	0.39	ng	-0.01
11) 2-Methylnaphthalene-d10	11.70	152	2497	0.47	ng	-0.01
14) 2,4,6-Tribromophenol	15.51	330	691	0.46	ng	-0.01
15) 2-Fluorobiphenyl	12.61	172	4485	0.33	ng	-0.01
25) Fluoranthene-d10	18.79	212	44867	0.49	ng	-0.01
29) Terphenyl-d14	19.41	244	10975	0.62	ng	-0.01
Target Compounds						
9) Naphthalene	10.17	128	42053	1.207	ng	# 90
12) 2-Methylnaphthalene	11.78	142	115909	21.323	ng	# 98
17) Acenaphthene	14.06	154	1408	0.137	ng	# 47
18) Fluorene	15.05	166	5364	0.401	ng	# 78
23) Phenanthrene	16.79	178	2185	0.128	ng	# 92
32) Bis(2-ethylhexyl)phthalate	20.89	149	6127	0.160	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092018\
Data File : BE097604.D
Acq On : 21 Sep 2018 10:28
Operator : SJ/JU
Sample : J4892-10
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PR-MW-03-091218

Quant Time: Sep 21 13:43:38 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 20 10:54:43 2018
Response via : Initial Calibration

