

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092018\
 Data File : BE097590.D
 Acq On : 21 Sep 2018 1:49
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Sep 21 06:59:38 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 20 10:54:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.707	0.682	3.5	92	0.00
3	n-Nitrosodimethylamine	0.624	0.488	21.8	93	0.00
4 S	2-Fluorophenol	1.199	1.125	6.2	99	0.00
5 S	Phenol-d6	1.532	1.389	9.3	94	0.00
6	bis(2-Chloroethyl)ether	1.421	1.293	9.0	97	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.309	0.275	11.0	103	0.00
9	Naphthalene	3.602	3.602	0.0	110	0.00
10	Hexachlorobutadiene	0.166	0.165	0.6	107	0.00
11 SURR	2-Methylnaphthalene-d10	0.551	0.600	-8.9	122	0.00
12	2-Methylnaphthalene	0.562	0.599	-6.6	120	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	134	0.00
14 S	2,4,6-Tribromophenol	0.166	0.171	-3.0	175#	0.00
15 S	2-Fluorobiphenyl	1.413	1.274	9.8	130	0.00
16	Acenaphthylene	1.581	1.557	1.5	148	0.00
17	Acenaphthene	1.060	1.027	3.1	139	0.00
18	Fluorene	1.382	1.419	-2.7	151#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	148	0.00
20	4-Bromophenyl-phenylether	0.180	0.173	3.9	154#	0.00
21	Hexachlorobenzene	0.198	0.203	-2.5	160#	0.00
22	Pentachlorophenol	0.069	0.061	11.6	162#	0.00
23	Phenanthrene	0.950	0.907	4.5	155#	0.00
24	Anthracene	0.839	0.776	7.5	153#	0.00
25 SURR	Fluoranthene-d10	5.121	4.229	17.4	133	0.00
26	Fluoranthene	1.318	1.103	16.3	134	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
28	Pvrene	0.910	1.037	-14.0	129	0.00
29 S	Terphenyl-d14	0.732	0.784	-7.1	122	0.00
30	Benzo(a)anthracene	1.007	0.949	5.8	110	0.00
31	Chrysene	1.080	1.079	0.1	109	0.00
32	Bis(2-ethylhexyl)phthalate	2.006	1.346	32.9#	95	0.00
33	Indeno(1,2,3-cd)pyrene	1.301	1.283	1.4	106	0.00
34 I	Perylene-d12	1.000	1.000	0.0	100	0.00
35	Benzo(b)fluoranthene	1.026	0.957	6.7	104	0.00
36	Benzo(k)fluoranthene	1.171	1.144	2.3	109	0.00
37 C	Benzo(a)pyrene	1.053	0.964	8.5	106	0.00
38	Dibenzo(a,h)anthracene	1.075	1.059	1.5	107	0.00
39	Benzo(a,h,i)perylene	1.046	1.021	2.4	105	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092018\
Data File : BE097590.D
Acq On : 21 Sep 2018 1:49
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 21 06:59:38 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 20 10:54:43 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			