

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092018\
 Data File : BE097607.D
 Acq On : 21 Sep 2018 12:19
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 21 13:48:16 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 20 10:54:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	127	-0.01
2	1,4-Dioxane	0.707	0.596	15.7	105	0.00
3	n-Nitrosodimethylamine	0.624	0.459	26.4#	115	0.00
4 S	2-Fluorophenol	1.199	1.090	9.1	125	0.00
5 S	Phenol-d6	1.532	1.677	-9.5	149	-0.01
6	bis(2-Chloroethyl)ether	1.421	1.343	5.5	131	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	130	-0.01
8 S	Nitrobenzene-d5	0.309	0.279	9.7	133	0.00
9	Naphthalene	3.602	3.659	-1.6	142	0.00
10	Hexachlorobutadiene	0.166	0.165	0.6	137	0.00
11 SURR	2-Methylnaphthalene-d10	0.551	0.604	-9.6	156#	0.00
12	2-Methylnaphthalene	0.562	0.602	-7.1	153#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	168#	0.00
14 S	2,4,6-Tribromophenol	0.166	0.180	-8.4	231#	0.00
15 S	2-Fluorobiphenyl	1.413	1.286	9.0	165#	0.00
16	Acenaphthylene	1.581	1.589	-0.5	190#	0.00
17	Acenaphthene	1.060	1.044	1.5	177#	0.00
18	Fluorene	1.382	1.439	-4.1	192#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	187#	0.00
20	4-Bromophenyl-phenylether	0.180	0.174	3.3	197#	0.00
21	Hexachlorobenzene	0.198	0.202	-2.0	203#	0.00
22	Pentachlorophenol	0.069	0.052	24.6	175#	0.00
23	Phenanthrene	0.950	0.929	2.2	201#	0.00
24	Anthracene	0.839	0.786	6.3	197#	0.00
25 SURR	Fluoranthene-d10	5.121	4.224	17.5	168#	0.00
26	Fluoranthene	1.318	1.102	16.4	170#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	136	0.00
28	Pvrene	0.910	1.000	-9.9	163#	0.00
29 S	Terphenyl-d14	0.732	0.752	-2.7	153#	0.00
30	Benzo(a)anthracene	1.007	0.945	6.2	143	0.00
31	Chrysene	1.080	1.050	2.8	139	0.00
32	Bis(2-ethylhexyl)phthalate	2.006	1.352	32.6#	125	-0.01
33	Indeno(1,2,3-cd)pyrene	1.301	1.271	2.3	138	0.00
34 I	Perylene-d12	1.000	1.000	0.0	128	0.00
35	Benzo(b)fluoranthene	1.026	0.934	9.0	130	0.00
36	Benzo(k)fluoranthene	1.171	1.148	2.0	140	0.00
37 C	Benzo(a)pyrene	1.053	0.952	9.6	133	0.00
38	Dibenzo(a,h)anthracene	1.075	1.067	0.7	138	0.00
39	Benzo(a,h,i)perylene	1.046	1.028	1.7	136	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092018\
Data File : BE097607.D
Acq On : 21 Sep 2018 12:19
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 21 13:48:16 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 20 10:54:43 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			