

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092018\
 Data File : BE097607.D
 Acq On : 21 Sep 2018 12:19
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 21 13:48:16 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 20 10:54:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	127	-0.01
2	1,4-Dioxane	0.400	0.343	14.2	105	0.00
3	n-Nitrosodimethylamine	0.400	0.320	20.0	115	0.00
4 S	2-Fluorophenol	0.400	0.364	9.0	125	0.00
5 S	Phenol-d6	0.400	0.438	-9.5	149	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.378	5.5	131	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	130	-0.01
8 S	Nitrobenzene-d5	0.400	0.362	9.5	133	0.00
9	Naphthalene	0.400	0.406	-1.5	142	0.00
10	Hexachlorobutadiene	0.400	0.398	0.5	137	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.439	-9.7	156	0.00
12	2-Methylnaphthalene	0.400	0.428	-7.0	153	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	168	0.00
14 S	2,4,6-Tribromophenol	0.400	0.461	-15.3	231	0.00
15 S	2-Fluorobiphenyl	0.400	0.364	9.0	165	0.00
16	Acenaphthylene	0.400	0.402	-0.5	190	0.00
17	Acenaphthene	0.400	0.394	1.5	177	0.00
18	Fluorene	0.400	0.417	-4.2	192	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	187	0.00
20	4-Bromophenyl-phenylether	0.400	0.386	3.5	197	0.00
21	Hexachlorobenzene	0.400	0.408	-2.0	203	0.00
22	Pentachlorophenol	0.400	0.362	9.5	175	0.00
23	Phenanthrene	0.400	0.391	2.3	201	0.00
24	Anthracene	0.400	0.375	6.3	197	0.00
25 SURR	Fluoranthene-d10	0.400	0.330	17.5	168	0.00
26	Fluoranthene	0.400	0.334	16.5	170	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	136	0.00
28	Pvrene	0.400	0.440	-10.0	163	0.00
29 S	Terphenyl-d14	0.400	0.411	-2.7	153	0.00
30	Benzo(a)anthracene	0.400	0.375	6.3	143	0.00
31	Chrysene	0.400	0.389	2.8	139	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.330	17.5	125	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.391	2.3	138	0.00
34 I	Perylene-d12	0.400	0.400	0.0	128	0.00
35	Benzo(b)fluoranthene	0.400	0.364	9.0	130	0.00
36	Benzo(k)fluoranthene	0.400	0.392	2.0	140	0.00
37 C	Benzo(a)pyrene	0.400	0.384	4.0	133	0.00
38	Dibenzo(a,h)anthracene	0.400	0.397	0.8	138	0.00
39	Benzo(a,h,i)perylene	0.400	0.393	1.8	136	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092018\
Data File : BE097607.D
Acq On : 21 Sep 2018 12:19
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 21 13:48:16 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 20 10:54:43 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				