

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE092319\
 Data File : BE100538.D
 Acq On : 23 Sep 2019 13:45
 Operator : HP/JU
 Sample : PB123105BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Quant Time: Sep 23 14:22:54 2019

Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092019.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Fri Sep 20 18:26:17 2019

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.86	152	5938	0.40	ng	0.00
7) Naphthalene-d8	10.65	136	27553	0.40	ng	0.00
13) Acenaphthene-d10	14.48	164	15400	0.40	ng	0.00
19) Phenanthrene-d10	17.20	188	33231	0.40	ng	0.00
27) Chrysene-d12	21.36	240	36060	0.40	ng	0.00
34) Perylene-d12	23.83	264	39853	0.40	ng	0.00


System Monitoring Compounds

4) 2-Fluorophenol	5.43	112	5197	0.37	ng	0.00
5) Phenol-d6	7.02	99	7059	0.37	ng	0.00
8) Nitrobenzene-d5	9.00	82	5844	0.37	ng	0.00
11) 2-Methylnaphthalene-d10	12.24	152	15485	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.96	330	880	0.35	ng	0.00
15) 2-Fluorobiphenyl	13.12	172	21869	0.40	ng	0.00
25) Fluoranthene-d10	19.23	212	147599	0.37	ng	0.00
29) Terphenyl-d14	19.82	244	34489	0.40	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.35	88	3492	0.408	ng	98
3) n-Nitrosodimethylamine	3.67	42	1770	0.352	ng	# 86
6) bis(2-Chloroethyl)ether	7.29	93	7368	0.383	ng	96
9) Naphthalene	10.69	128	111343	0.390	ng	100
10) Hexachlorobutadiene	10.99	225	3602	0.394	ng	98
12) 2-Methylnaphthalene	12.31	142	17528	0.386	ng	99
16) Acenaphthylene	14.20	152	23201	0.366	ng	100
17) Acenaphthene	14.54	154	16545	0.386	ng	100
18) Fluorene	15.51	166	20505	0.383	ng	99
20) 4-Bromophenyl-phenylether	16.40	248	6104	0.391	ng	97
21) Hexachlorobenzene	16.52	284	5311	0.390	ng	99
22) Pentachlorophenol	16.87	266	726	0.399	ng	96
23) Phenanthrene	17.24	178	36304	0.392	ng	99
24) Anthracene	17.34	178	27971	0.359	ng	99
26) Fluoranthene	19.25	202	42089	0.377	ng	100
28) Pyrene	19.62	202	42212	0.407	ng	100
30) Benzo(a)anthracene	21.34	228	35866	0.372	ng	100
31) Chrysene	21.40	228	42601	0.388	ng	100
32) Bis(2-ethylhexyl)phthalate	21.28	149	34720	0.334	ng	100
33) Indeno(1,2,3-cd)pyrene	26.45	276	45468	0.373	ng	100
35) Benzo(b)fluoranthene	23.08	252	38662	0.359	ng	100
36) Benzo(k)fluoranthene	23.13	252	46251	0.371	ng	99
37) Benzo(a)pyrene	23.73	252	35530	0.353	ng	99
38) Dibenzo(a,h)anthracene	26.47	278	37556	0.358	ng	99
39) Benzo(g,h,i)perylene	27.24	276	40126	0.367	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1

1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.86 min Scan# 434
Delta R.T. -0.00 min
Lab File: BE100538.D
Acq: 23 Sep 2019 13:45

Tgt Ion: 152 Resp: 5938
Ion Ratio Lower Upper
152 100
150 152.4 121.6 182.4
115 62.0 48.2 72.4




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 0.408 ng
RT: 3.35 min Scan# 42
Delta R.T. 0.01 min
Lab File: BE100538.D
Acq: 23 Sep 2019 13:45

Tgt Ion: 88 Resp: 3492
Ion Ratio Lower Upper
88 100
58 65.3 53.3 79.9
43 21.8 16.9 25.3



Abundance Ion 88.00 (87.70 to 88.70): BE1
Ion 58.00 (57.70 to 58.70): BE1
Ion 43.00 (42.70 to 43.70): BE1









#3

n-Nitrosodimethylamine
 Concen: 0.352 ng
 RT: 3.67 min Scan# 70
 Delta R.T. -0.00 min
 Lab File: BE100538.D
 Acq: 23 Sep 2019 13:45

Tgt Ion	Ratio	Lower	Upper
42	100		
74	148.1	133.8	200.6
44	1.6	6.4	9.6#




Abundance Ion 42.00 (41.70 to 42.70): BE1
 Ion 74.00 (73.70 to 74.70): BE1
 Ion 44.00 (43.70 to 44.70): BE1





#4


2-Fluorophenol
 Concen: 0.374 ng
 RT: 5.43 min Scan# 223
 Delta R.T. -0.00 min
 Lab File: BE100538.D
 Acq: 23 Sep 2019 13:45

Tgt Ion	Ratio	Lower	Upper
112	100		
64	55.1	44.1	66.1
63	109.8	86.5	129.7



Abundance Ion 112.00 (111.70 to 112.70): E
 Ion 64.00 (63.70 to 64.70): BE1
 Ion 63.00 (62.70 to 63.70): BE1





#5

Phenol-d6

Concen: 0.371 ng

RT: 7.02 min Scan# 361

Delta R.T. -0.00 min

Lab File: BE100538.D

Acq: 23 Sep 2019 13:45


Tgt Ion: 99 Resp: 7059

Ion Ratio Lower Upper




99 100

42 15.6 13.0 19.6

71 29.4 23.4 35.0



Abundance Ion 99.00 (98.70 to 99.70): BE1
Ion 42.00 (41.70 to 42.70): BE1
Ion 71.00 (70.70 to 71.70): BE1



#6

bis(2-Chloroethyl)ether

Concen: 0.383 ng

RT: 7.29 min Scan# 384

Delta R.T. 0.01 min

Lab File: BE100538.D

Acq: 23 Sep 2019 13:45


Tgt Ion: 93 Resp: 7368

Ion Ratio Lower Upper



93 100


63 273.7 213.4 320.0

95 32.6 23.7 35.5



Abundance Ion 93.00 (92.70 to 93.70): BE1
Ion 63.00 (62.70 to 63.70): BE1
Ion 95.00 (94.70 to 95.70): BE1





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.65 min Scan# 653

Delta R.T. -0.00 min

Lab File: BE100538.D

Acq: 23 Sep 2019 13:45

Tgt Ion: 136 Resp: 27553


Ion Ratio Lower Upper

136 100

137 11.4 8.6 13.0

54 15.9 12.4 18.6

68 4.0 3.3 4.9






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE1

Ion 68.00 (67.70 to 68.70): BE1



#8

Nitrobenzene-d5

Concen: 0.368 ng

RT: 9.00 min Scan# 526

Delta R.T. -0.00 min

Lab File: BE100538.D

Acq: 23 Sep 2019 13:45


Tgt Ion: 82 Resp: 5844

Ion Ratio Lower Upper

82 100

128 159.9 129.4 194.2



54 179.9 143.3 214.9




Abundance Ion 82.00 (81.70 to 82.70): BE1

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE1





#9

Naphthalene

Concen: 0.390 ng

RT: 10.69 min Scan# 656

Delta R.T. -0.00 min

Lab File: BE100538.D

Acq: 23 Sep 2019 13:45


Tgt Ion:128 Resp: 111343

Ion Ratio Lower Upper

128 100

129 2.8 2.2 3.4




127 3.3 2.7 4.1



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#10

Hexachlorobutadiene

Concen: 0.394 ng

RT: 10.99 min Scan# 679

Delta R.T. -0.00 min

Lab File: BE100538.D

Acq: 23 Sep 2019 13:45


Tgt Ion:225 Resp: 3602

Ion Ratio Lower Upper

225 100

223 63.2 49.7 74.5



227 62.7 51.4 77.2




Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E





Ion 227.00 (226.70 to 227.70): E








#11
 2-Methylnaphthalene-d10
 Concen: 0.376 ng
 RT: 12.24 min Scan# 765
 Delta R.T. -0.00 min
 Lab File: BE100538.D
 Acq: 23 Sep 2019 13:45


Tgt Ion:152 Resp: 15485
 Ion Ratio Lower Upper
 152 100
 151 19.3 15.3 22.9



#12
 2-Methylnaphthalene
 Concen: 0.386 ng
 RT: 12.31 min Scan# 770
 Delta R.T. -0.00 min
 Lab File: BE100538.D
 Acq: 23 Sep 2019 13:45





Tgt Ion:142 Resp: 17528
 Ion Ratio Lower Upper
 142 100
 141 90.0 72.5 108.7
 115 33.1 27.2 40.8








#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.48 min Scan# 921
Delta R.T. -0.00 min
Lab File: BE100538.D
Acq: 23 Sep 2019 13:45


Tgt Ion:164 Resp: 15400
Ion Ratio Lower Upper
164 100
162 104.3 82.1 123.1
160 47.6 37.6 56.4



#14
2,4,6-Tribromophenol
Concen: 0.353 ng
RT: 15.96 min Scan# 1025
Delta R.T. -0.00 min
Lab File: BE100538.D
Acq: 23 Sep 2019 13:45





Tgt Ion:330 Resp: 880
Ion Ratio Lower Upper
330 100
332 92.5 75.5 113.3
141 52.2 42.3 63.5








#15
 2-Fluorobiphenyl
 Concen: 0.398 ng
 RT: 13.12 min Scan# 826
 Delta R.T. -0.00 min
 Lab File: BE100538.D
 Acq: 23 Sep 2019 13:45


Tgt Ion:	172	Resp:	21869
Ion	Ratio	Lower	Upper
172	100		
171	35.3	28.8	43.2
170	24.4	19.4	29.0



#16
 Acenaphthylene
 Concen: 0.366 ng
 RT: 14.20 min Scan# 901
 Delta R.T. -0.00 min
 Lab File: BE100538.D
 Acq: 23 Sep 2019 13:45

Tgt Ion:	152	Resp:	23201
Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.9	23.9
153	13.4	10.7	16.1





#17

Acenaphthene

Concen: 0.386 ng

RT: 14.54 min Scan# 925

Delta R.T. -0.00 min

Lab File: BE100538.D

Acq: 23 Sep 2019 13:45





Tgt Ion:154 Resp: 16545

Ion Ratio Lower Upper

154 100

153 116.9 93.7 140.5

152 58.7 47.3 70.9



#18

Fluorene

Concen: 0.383 ng

RT: 15.51 min Scan# 992

Delta R.T. -0.00 min

Lab File: BE100538.D

Acq: 23 Sep 2019 13:45




Tgt Ion:166 Resp: 20505


Ion Ratio Lower Upper

166 100

165 99.5 80.1 120.1





167 13.6 10.9 16.3








#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.20 min Scan# 1118
Delta R.T. -0.00 min
Lab File: BE100538.D
Acq: 23 Sep 2019 13:45


Tgt Ion:188 Resp: 33231
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.0 8.0 12.0



#20
4-Bromophenyl-phenylether
Concen: 0.391 ng
RT: 16.40 min Scan# 1058
Delta R.T. -0.00 min
Lab File: BE100538.D
Acq: 23 Sep 2019 13:45

Tgt Ion:248 Resp: 6104
Ion Ratio Lower Upper
248 100
250 87.6 73.5 110.3
141 95.6 77.0 115.4





#21

Hexachlorobenzene

Concen: 0.390 ng


RT: 16.52 min Scan# 1067

Delta R.T. -0.00 min




Lab File: BE100538.D

Acq: 23 Sep 2019 13:45

Tgt Ion:	284	Resp:	5311
Ion	Ratio	Lower	Upper
284	100		
142	47.6	39.0	58.6
249	35.2	28.2	42.2



Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E



#22

Pentachlorophenol

Concen: 0.399 ng


RT: 16.87 min Scan# 1093

Delta R.T. -0.01 min



Lab File: BE100538.D


Acq: 23 Sep 2019 13:45

Tgt Ion:	266	Resp:	726
Ion	Ratio	Lower	Upper
266	100		
264	70.6	57.6	86.4
268	74.4	55.1	82.7







Abundance Ion 266.00 (265.70 to 266.70): E
Ion 264.00 (263.70 to 264.70): E
Ion 268.00 (267.70 to 268.70): E








#23
Phenanthrene
Concen: 0.392 ng
RT: 17.24 min Scan# 1121
Delta R.T. -0.00 min
Lab File: BE100538.D
Acq: 23 Sep 2019 13:45


Tgt Ion:178 Resp: 36304
Ion Ratio Lower Upper
178 100
176 18.7 15.4 23.2
179 15.5 12.6 18.8



#24
Anthracene
Concen: 0.359 ng
RT: 17.34 min Scan# 1128
Delta R.T. -0.00 min
Lab File: BE100538.D
Acq: 23 Sep 2019 13:45





Tgt Ion:178 Resp: 27971
Ion Ratio Lower Upper
178 100
176 18.7 15.2 22.8
179 15.1 12.2 18.4








#25
 Fluoranthene-d10
 Concen: 0.373 ng
 RT: 19.23 min Scan# 1385
 Delta R.T. 0.00 min
 Lab File: BE100538.D
 Acq: 23 Sep 2019 13:45

Tgt Ion: 212 Resp: 147599
 Ion Ratio Lower Upper
 212 100
 106 3.0 2.3 3.5
 104 1.7 1.3 1.9



#26
 Fluoranthene
 Concen: 0.377 ng
 RT: 19.25 min Scan# 1390
 Delta R.T. -0.00 min
 Lab File: BE100538.D
 Acq: 23 Sep 2019 13:45


Tgt Ion: 202 Resp: 42089
 Ion Ratio Lower Upper
 202 100
 101 11.1 8.8 13.2
 203 17.2 13.7 20.5








#27
Chrysene-d12
Concen: 0.400 ng
RT: 21.36 min Scan# 1629
Delta R.T. -0.00 min
Lab File: BE100538.D
Acq: 23 Sep 2019 13:45

Tgt Ion: 240 Resp: 36060
Ion Ratio Lower Upper
240 100
120 6.5 4.5 6.7
236 27.0 20.6 30.8




Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E






#28
Pyrene
Concen: 0.407 ng
RT: 19.62 min Scan# 1453
Delta R.T. 0.00 min
Lab File: BE100538.D
Acq: 23 Sep 2019 13:45

Tgt Ion: 202 Resp: 42212
Ion Ratio Lower Upper
202 100
200 20.8 16.6 25.0
203 17.7 14.0 21.0







Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E








#29
 Terphenyl-d14
 Concen: 0.404 ng
 RT: 19.82 min Scan# 1489
 Delta R.T. 0.00 min
 Lab File: BE100538.D
 Acq: 23 Sep 2019 13:45


Tgt Ion: 244 Resp: 34489
 Ion Ratio Lower Upper
 244 100
 212 27.5 22.2 33.2
 122 10.0 8.3 12.5



#30
 Benzo(a)anthracene
 Concen: 0.372 ng
 RT: 21.34 min Scan# 1627
 Delta R.T. -0.00 min
 Lab File: BE100538.D
 Acq: 23 Sep 2019 13:45


Tgt Ion: 228 Resp: 35866
 Ion Ratio Lower Upper
 228 100
 226 28.3 22.5 33.7
 229 19.4 15.4 23.2








#31
Chrysene
Concen: 0.388 ng
RT: 21.40 min Scan# 1632
Delta R.T. -0.00 min
Lab File: BE100538.D
Acq: 23 Sep 2019 13:45

Tgt Ion: 228 Resp: 42601
Ion Ratio Lower Upper
228 100
226 28.9 23.3 34.9
229 20.3 16.3 24.5




Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E






#32
Bis(2-ethylhexyl)phthalate
Concen: 0.334 ng
RT: 21.28 min Scan# 1622
Delta R.T. -0.00 min
Lab File: BE100538.D
Acq: 23 Sep 2019 13:45

Tgt Ion: 149 Resp: 34720
Ion Ratio Lower Upper
149 100
167 6.6 5.2 7.8
279 1.1 1.0 1.4



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E





#33

Indeno(1,2,3-cd)pyrene

Concen: 0.373 ng

RT: 26.45 min Scan# 2918

Delta R.T. -0.00 min

Lab File: BE100538.D

Acq: 23 Sep 2019 13:45


Tgt Ion: 276 Resp: 45468

Ion Ratio Lower Upper

276 100

138 22.3 17.7 26.5




277 25.1 19.9 29.9



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#34

Perylene-d12

Concen: 0.400 ng

RT: 23.83 min Scan# 2189

Delta R.T. -0.00 min

Lab File: BE100538.D

Acq: 23 Sep 2019 13:45


Tgt Ion: 264 Resp: 39853

Ion Ratio Lower Upper

264 100

260 24.2 19.6 29.4



265 26.5 21.8 32.8



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E





Ion 265.00 (264.70 to 265.70): E








#35
Benzo(b)fluoranthene
Concen: 0.359 ng
RT: 23.08 min Scan# 1977
Delta R.T. 0.00 min
Lab File: BE100538.D
Acq: 23 Sep 2019 13:45

Tgt Ion: 252 Resp: 38662
Ion Ratio Lower Upper
252 100
253 22.3 18.1 27.1
125 10.1 8.0 12.0



#36
Benzo(k)fluoranthene
Concen: 0.371 ng
RT: 23.13 min Scan# 1991
Delta R.T. 0.00 min
Lab File: BE100538.D
Acq: 23 Sep 2019 13:45


Tgt Ion: 252 Resp: 46251
Ion Ratio Lower Upper
252 100
253 22.4 17.9 26.9
125 11.2 8.5 12.7








#37
Benzo(a)pyrene
Concen: 0.353 ng
RT: 23.73 min Scan# 2159
Delta R.T. 0.00 min
Lab File: BE100538.D
Acq: 23 Sep 2019 13:45

Tgt Ion: 252 Resp: 35530
Ion Ratio Lower Upper
252 100
253 22.9 18.1 27.1
125 11.7 9.2 13.8




Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E






#38
Dibenzo(a,h)anthracene
Concen: 0.358 ng
RT: 26.47 min Scan# 2925
Delta R.T. -0.00 min
Lab File: BE100538.D
Acq: 23 Sep 2019 13:45

Tgt Ion: 278 Resp: 37556
Ion Ratio Lower Upper
278 100
139 14.2 11.6 17.4
279 23.7 19.7 29.5



Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E





#39

Benzo(g,h,i)perylene

Concen: 0.367 ng

RT: 27.24 min Scan# 3140

Delta R.T. -0.00 min

Lab File: BE100538.D

Acq: 23 Sep 2019 13:45


Tgt Ion: 276 Resp: 40126

Ion Ratio Lower Upper

276 100

277 24.1 19.2 28.8

138 20.1 15.7 23.5



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

