

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092418\
 Data File : BE097650.D
 Acq On : 24 Sep 2018 19:13
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE092418

Quant Time: Sep 25 11:59:44 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 24 18:40:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	113	0.00
2	1,4-Dioxane	0.400	0.387	3.3	118	0.00
3	n-Nitrosodimethylamine	0.400	0.403	-0.8	129	0.00
4 S	2-Fluorophenol	0.400	0.396	1.0	127	0.00
5 S	Phenol-d6	0.400	0.430	-7.5	139	0.00
6	bis(2-Chloroethyl)ether	0.400	0.423	-5.7	119	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	112	0.00
8 S	Nitrobenzene-d5	0.400	0.427	-6.7	135	0.00
9	Naphthalene	0.400	0.417	-4.2	125	-0.01
10	Hexachlorobutadiene	0.400	0.410	-2.5	125	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.405	-1.3	121	0.00
12	2-Methylnaphthalene	0.400	0.426	-6.5	128	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.400	0.397	0.8	130	0.00
15 S	2-Fluorobiphenyl	0.400	0.426	-6.5	125	-0.01
16	Acenaphthylene	0.400	0.400	0.0	123	0.00
17	Acenaphthene	0.400	0.420	-5.0	125	0.00
18	Fluorene	0.400	0.404	-1.0	121	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	109	0.00
20	4-Bromophenyl-phenylether	0.400	0.414	-3.5	122	0.00
21	Hexachlorobenzene	0.400	0.408	-2.0	119	0.00
22	Pentachlorophenol	0.400	0.419	-4.7	132	-0.01
23	Phenanthrene	0.400	0.408	-2.0	120	0.00
24	Anthracene	0.400	0.408	-2.0	125	0.00
25 SURR	Fluoranthene-d10	0.400	0.397	0.8	121	0.00
26	Fluoranthene	0.400	0.403	-0.8	122	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	110	0.00
28	Pvrene	0.400	0.435	-8.7	129	0.00
29 S	Terphenyl-d14	0.400	0.423	-5.7	127	0.00
30	Benzo(a)anthracene	0.400	0.405	-1.3	128	0.00
31	Chrysene	0.400	0.422	-5.5	119	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.419	-4.7	132	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.443	-10.7	130	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	112	0.00
35	Benzo(b)fluoranthene	0.400	0.398	0.5	120	0.00
36	Benzo(k)fluoranthene	0.400	0.416	-4.0	125	0.00
37 C	Benzo(a)pyrene	0.400	0.406	-1.5	126	0.00
38	Dibenzo(a,h)anthracene	0.400	0.422	-5.5	133	-0.01
39	Benzo(a,h,i)perylene	0.400	0.410	-2.5	125	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092418\
Data File : BE097650.D
Acq On : 24 Sep 2018 19:13
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE092418

Quant Time: Sep 25 11:59:44 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 24 18:40:34 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				