

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092418\
 Data File : BE097650.D
 Acq On : 24 Sep 2018 19:13
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE092418

Quant Time: Sep 25 11:59:44 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 24 18:40:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
2	1,4-Dioxane	0.716	0.669	6.6	118	0.00
3	n-Nitrosodimethylamine	0.421	0.451	-7.1	129	0.00
4 S	2-Fluorophenol	1.053	1.043	0.9	127	0.00
5 S	Phenol-d6	1.487	1.596	-7.3	139	0.00
6	bis(2-Chloroethyl)ether	1.204	1.312	-9.0	119	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
8 S	Nitrobenzene-d5	0.232	0.264	-13.8	135	0.00
9	Naphthalene	3.636	3.786	-4.1	125	-0.01
10	Hexachlorobutadiene	0.165	0.169	-2.4	125	0.00
11 SURR	2-Methylnaphthalene-d10	0.592	0.600	-1.4	121	0.00
12	2-Methylnaphthalene	0.576	0.614	-6.6	128	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.155	0.159	-2.6	130	0.00
15 S	2-Fluorobiphenyl	1.305	1.392	-6.7	125	-0.01
16	Acenaphthylene	1.608	1.609	-0.1	123	0.00
17	Acenaphthene	1.050	1.103	-5.0	125	0.00
18	Fluorene	1.477	1.492	-1.0	121	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
20	4-Bromophenyl-phenylether	0.177	0.183	-3.4	122	0.00
21	Hexachlorobenzene	0.204	0.209	-2.5	119	0.00
22	Pentachlorophenol	0.049	0.041	16.3	132	-0.01
23	Phenanthrene	0.946	0.966	-2.1	120	0.00
24	Anthracene	0.786	0.802	-2.0	125	0.00
25 SURR	Fluoranthene-d10	4.361	4.326	0.8	121	0.00
26	Fluoranthene	1.135	1.143	-0.7	122	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
28	Pvrene	1.044	1.136	-8.8	129	0.00
29 S	Terphenyl-d14	0.787	0.834	-6.0	127	0.00
30	Benzo(a)anthracene	0.945	0.958	-1.4	128	0.00
31	Chrysene	1.106	1.166	-5.4	119	0.00
32	Bis(2-ethylhexyl)phthalate	1.080	1.130	-4.6	132	0.00
33	Indeno(1,2,3-cd)pyrene	1.165	1.290	-10.7	130	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	112	0.00
35	Benzo(b)fluoranthene	1.023	1.017	0.6	120	0.00
36	Benzo(k)fluoranthene	1.211	1.258	-3.9	125	0.00
37 C	Benzo(a)pyrene	1.010	1.026	-1.6	126	0.00
38	Dibenzo(a,h)anthracene	1.062	1.121	-5.6	133	-0.01
39	Benzo(a,h,i)perylene	1.056	1.082	-2.5	125	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092418\
Data File : BE097650.D
Acq On : 24 Sep 2018 19:13
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE092418

Quant Time: Sep 25 11:59:44 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 24 18:40:34 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			