

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092519\
 Data File : BE100563.D
 Acq On : 25 Sep 2019 19:04
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 18:15:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 20 18:26:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
2	1,4-Dioxane	0.577	0.521	9.7	74	0.00
3	n-Nitrosodimethylamine	0.339	0.369	-8.8	118	-0.01
4 S	2-Fluorophenol	0.936	1.217	-30.0#	116	0.01
5 S	Phenol-d6	1.282	1.753	-36.7#	125	0.00
6	bis(2-Chloroethyl)ether	1.297	1.317	-1.5	88	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
8 S	Nitrobenzene-d5	0.231	0.295	-27.7#	132	0.00
9	Naphthalene	4.147	4.047	2.4	92	0.00
10	Hexachlorobutadiene	0.133	0.122	8.3	86	0.00
11 SURR	2-Methylnaphthalene-d10	0.598	0.607	-1.5	99	0.00
12	2-Methylnaphthalene	0.660	0.674	-2.1	100	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
14 S	2,4,6-Tribromophenol	0.065	0.134	-106.2#	212#	0.00
15 S	2-Fluorobiphenyl	1.428	1.375	3.7	92	0.00
16	Acenaphthylene	1.647	3.175	-92.8#	198#	0.01
17	Acenaphthene	1.112	1.212	-9.0	108	0.01
18	Fluorene	1.391	1.511	-8.6	108	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	84	0.00
20	4-Bromophenyl-phenylether	0.188	0.200	-6.4	93	0.00
21	Hexachlorobenzene	0.164	0.157	4.3	85	0.01
22	Pentachlorophenol	0.037	0.073	-97.3#	277#	-0.01
23	Phenanthrene	1.115	1.292	-15.9	101	0.00
24	Anthracene	0.938	1.147	-22.3	114	0.00
25 SURR	Fluoranthene-d10	4.758	4.539	4.6	84	0.00
26	Fluoranthene	1.344	1.482	-10.3	98	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	75	0.00
28	Pvrene	1.151	1.638	-42.3#	111	0.00
29 S	Terphenyl-d14	0.947	1.045	-10.3	86	0.00
30	Benzo(a)anthracene	1.068	1.417	-32.7#	105	0.01
31	Chrysene	1.217	1.299	-6.7	84	0.00
32	Bis(2-ethylhexyl)phthalate	1.151	3.369	-192.7#	244#	0.00
33	Indeno(1,2,3-cd)pyrene	1.353	1.566	-15.7	90	0.01
34 I	Perylene-d12	1.000	1.000	0.0	82	0.00
35	Benzo(b)fluoranthene	1.082	1.167	-7.9	98	0.01
36	Benzo(k)fluoranthene	1.252	1.133	9.5	80	0.00
37 C	Benzo(a)pyrene	1.010	1.144	-13.3	100	0.00
38	Dibenzo(a,h)anthracene	1.052	1.073	-2.0	91	0.00
39	Benzo(a,h,i)perylene	1.098	1.117	-1.7	87	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092519\
Data File : BE100563.D
Acq On : 25 Sep 2019 19:04
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 26 18:15:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 20 18:26:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			