

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE092818\
 Data File : BE097698.D
 Acq On : 28 Sep 2018 9:27
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 28 11:04:47 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 27 15:45:00 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.834	0.792	5.0	96	0.00
3	n-Nitrosodimethylamine	0.501	0.438	12.6	86	0.00
4 S	2-Fluorophenol	0.989	0.847	14.4	86	0.00
5 S	Phenol-d6	1.363	1.293	5.1	96	0.00
6	bis(2-Chloroethyl)ether	1.335	1.295	3.0	96	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.284	0.269	5.3	95	0.00
9	Naphthalene	3.951	3.824	3.2	98	0.00
10	Hexachlorobutadiene	0.185	0.179	3.2	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.622	0.597	4.0	100	0.00
12	2-Methylnaphthalene	0.604	0.578	4.3	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.156	0.127	18.6	96	0.00
15 S	2-Fluorobiphenyl	1.680	1.574	6.3	100	0.00
16	Acenaphthylene	1.707	1.554	9.0	99	0.00
17	Acenaphthene	1.149	1.098	4.4	99	0.00
18	Fluorene	1.468	1.369	6.7	98	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
20	4-Bromophenyl-phenylether	0.204	0.198	2.9	99	0.00
21	Hexachlorobenzene	0.229	0.230	-0.4	103	0.00
22	Pentachlorophenol	0.053	0.043	18.9	82	0.01
23	Phenanthrene	1.034	1.018	1.5	100	0.00
24	Anthracene	0.851	0.775	8.9	97	0.00
25 SURR	Fluoranthene-d10	4.850	4.756	1.9	101	0.00
26	Fluoranthene	1.284	1.252	2.5	100	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
28	Pyrene	1.041	1.009	3.1	101	0.00
29 S	Terphenyl-d14	0.864	0.811	6.1	98	0.00
30	Benzo(a)anthracene	0.989	0.930	6.0	101	0.00
31	Chrysene	1.238	1.192	3.7	100	0.00
32	Bis(2-ethylhexyl)phthalate	1.220	1.025	16.0	96	0.00
33	Indeno(1,2,3-cd)pyrene	1.307	1.315	-0.6	106	0.00
34 I	Perylene-d12	1.000	1.000	0.0	105	0.00
35	Benzo(b)fluoranthene	1.156	1.011	12.5	98	0.00
36	Benzo(k)fluoranthene	1.334	1.295	2.9	105	0.00
37 C	Benzo(a)pyrene	1.109	1.011	8.8	103	0.00
38	Dibenzo(a,h)anthracene	1.203	1.102	8.4	104	0.00
39	Benzo(g,h,i)perylene	1.220	1.089	10.7	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE092818\
Data File : BE097698.D
Acq On : 28 Sep 2018 9:27
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 28 11:04:47 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 27 15:45:00 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		