

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE092818\
 Data File : BE097704.D
 Acq On : 28 Sep 2018 15:01
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 28 15:46:07 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 27 15:45:00 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.834	0.771	7.6	101	0.00
3	n-Nitrosodimethylamine	0.501	0.494	1.4	105	0.00
4 S	2-Fluorophenol	0.989	0.893	9.7	98	0.00
5 S	Phenol-d6	1.363	1.110	18.6	90	0.00
6	bis(2-Chloroethyl)ether	1.335	1.369	-2.5	110	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.284	0.282	0.7	105	0.00
9	Naphthalene	3.951	3.802	3.8	102	0.00
10	Hexachlorobutadiene	0.185	0.178	3.8	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.622	0.596	4.2	104	0.00
12	2-Methylnaphthalene	0.604	0.580	4.0	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.156	0.130	16.7	103	0.00
15 S	2-Fluorobiphenyl	1.680	1.601	4.7	106	0.00
16	Acenaphthylene	1.707	1.600	6.3	107	0.00
17	Acenaphthene	1.149	1.092	5.0	102	0.00
18	Fluorene	1.468	1.394	5.0	104	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
20	4-Bromophenyl-phenylether	0.204	0.190	6.9	104	0.00
21	Hexachlorobenzene	0.229	0.216	5.7	107	0.00
22	Pentachlorophenol	0.053	0.045	15.1	94	0.00
23	Phenanthrene	1.034	0.983	4.9	107	0.00
24	Anthracene	0.851	0.779	8.5	107	0.00
25 SURR	Fluoranthene-d10	4.850	4.511	7.0	106	0.00
26	Fluoranthene	1.284	1.196	6.9	106	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
28	Pyrene	1.041	1.032	0.9	106	0.00
29 S	Terphenyl-d14	0.864	0.830	3.9	104	0.00
30	Benzo(a)anthracene	0.989	0.919	7.1	102	0.00
31	Chrysene	1.238	1.195	3.5	103	0.00
32	Bis(2-ethylhexyl)phthalate	1.220	1.017	16.6	98	0.00
33	Indeno(1,2,3-cd)pyrene	1.307	1.366	-4.5	114	0.00
34 I	Perylene-d12	1.000	1.000	0.0	111	0.00
35	Benzo(b)fluoranthene	1.156	1.001	13.4	102	0.00
36	Benzo(k)fluoranthene	1.334	1.305	2.2	112	0.00
37 C	Benzo(a)pyrene	1.109	1.019	8.1	110	0.00
38	Dibenzo(a,h)anthracene	1.203	1.141	5.2	113	0.00
39	Benzo(g,h,i)perylene	1.220	1.098	10.0	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE092818\
Data File : BE097704.D
Acq On : 28 Sep 2018 15:01
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 28 15:46:07 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 27 15:45:00 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			