

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE093019\
 Data File : BE100583.D
 Acq On : 30 Sep 2019 16:10
 Operator : JU
 Sample : K4839-08
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PT-ACIDS-WP

Quant Time: Oct 01 05:12:07 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 27 18:32:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.86	152	16073	0.40	ng	0.00
7) Naphthalene-d8	10.65	136	17799	0.40	ng	0.00
13) Acenaphthene-d10	14.48	164	9132	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	20261	0.40	ng	0.00
27) Chrysene-d12	21.37	240	24958	0.40	ng	0.00
34) Perylene-d12	23.85	264	30337	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.43	112	1878304	39.63	ng	0.00
5) Phenol-d6	7.03	99	2655462	41.27	ng	0.01
8) Nitrobenzene-d5	9.00	82	1503339	116.96	ng	0.00
11) 2-Methylnaphthalene-d10	12.25	152	375	0.01	ng	0.01
14) 2,4,6-Tribromophenol	15.97	330	563512	215.23	ng	0.01
15) 2-Fluorobiphenyl	13.13	172	3253445	102.92	ng	0.01
25) Fluoranthene-d10	19.24	212	2569	0.01	ng	0.00
29) Terphenyl-d14	19.83	244	4898205	79.43	ng	0.00
Target Compounds						
22) Pentachlorophenol	16.87	266	534382	126.543	ng	Qvalue 92
32) Bis(2-ethylhexyl)phthalate	21.29	149	7836	0.040	ng	# 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE093019\
Data File : BE100583.D
Acq On : 30 Sep 2019 16:10
Operator : JU
Sample : K4839-08
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PT-ACIDS-WP

Quant Time: Oct 01 05:12:07 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 27 18:32:20 2019
Response via : Initial Calibration

