

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE093019\
 Data File : BE100586.D
 Acq On : 30 Sep 2019 18:03
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 30 23:07:29 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 27 18:32:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.721	0.689	4.4	99	0.00
3	n-Nitrosodimethylamine	0.384	0.320	16.7	89	0.00
4 S	2-Fluorophenol	1.179	1.026	13.0	88	0.00
5 S	Phenol-d6	1.601	1.364	14.8	87	0.00
6	bis(2-Chloroethyl)ether	1.383	1.262	8.7	92	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	-0.01
8 S	Nitrobenzene-d5	0.289	0.244	15.6	86	0.00
9	Naphthalene	4.156	4.056	2.4	96	0.00
10	Hexachlorobutadiene	0.127	0.125	1.6	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.623	0.583	6.4	93	0.00
12	2-Methylnaphthalene	0.707	0.666	5.8	94	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.115	0.090	21.7	79	0.00
15 S	2-Fluorobiphenyl	1.385	1.392	-0.5	97	0.00
16	Acenaphthylene	1.782	1.648	7.5	91	0.00
17	Acenaphthene	1.145	1.101	3.8	94	-0.01
18	Fluorene	1.488	1.426	4.2	92	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
20	4-Bromophenyl-phenylether	0.186	0.175	5.9	96	0.00
21	Hexachlorobenzene	0.153	0.152	0.7	96	0.00
22	Pentachlorophenol	0.055	0.031	43.6#	71	0.00
23	Phenanthrene	1.128	1.125	0.3	95	0.00
24	Anthracene	1.021	0.926	9.3	92	0.00
25 SURR	Fluoranthene-d10	5.084	4.575	10.0	88	0.00
26	Fluoranthene	1.418	1.328	6.3	89	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	86	0.00
28	Pvrene	1.211	1.262	-4.2	87	0.00
29 S	Terphenyl-d14	0.988	0.967	2.1	87	0.00
30	Benzo(a)anthracene	1.178	1.112	5.6	82	0.00
31	Chrysene	1.201	1.210	-0.7	86	0.00
32	Bis(2-ethylhexyl)phthalate	3.127	1.927	38.4#	57	0.00
33	Indeno(1,2,3-cd)pyrene	1.481	1.440	2.8	83	0.00
34 I	Perylene-d12	1.000	1.000	0.0	86	0.00
35	Benzo(b)fluoranthene	1.094	1.002	8.4	83	0.00
36	Benzo(k)fluoranthene	1.193	1.128	5.4	86	0.00
37 C	Benzo(a)pyrene	1.045	0.960	8.1	82	0.00
38	Dibenzo(a,h)anthracene	1.065	0.988	7.2	84	0.00
39	Benzo(a,h,i)perylene	1.088	1.018	6.4	83	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE093019\
Data File : BE100586.D
Acq On : 30 Sep 2019 18:03
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 30 23:07:29 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 27 18:32:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			