

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE093019\
 Data File : BE100586.D
 Acq On : 30 Sep 2019 18:03
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 30 23:07:29 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 27 18:32:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	97	0.00
2	1,4-Dioxane	0.400	0.383	4.3	99	0.00
3	n-Nitrosodimethylamine	0.400	0.333	16.8	89	0.00
4 S	2-Fluorophenol	0.400	0.348	13.0	88	0.00
5 S	Phenol-d6	0.400	0.341	14.8	87	0.00
6	bis(2-Chloroethyl)ether	0.400	0.365	8.8	92	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	96	-0.01
8 S	Nitrobenzene-d5	0.400	0.337	15.8	86	0.00
9	Naphthalene	0.400	0.390	2.5	96	0.00
10	Hexachlorobutadiene	0.400	0.394	1.5	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.374	6.5	93	0.00
12	2-Methylnaphthalene	0.400	0.377	5.8	94	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.400	0.314	21.5	79	0.00
15 S	2-Fluorobiphenyl	0.400	0.402	-0.5	97	0.00
16	Acenaphthylene	0.400	0.370	7.5	91	0.00
17	Acenaphthene	0.400	0.385	3.8	94	-0.01
18	Fluorene	0.400	0.383	4.3	92	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	94	0.00
20	4-Bromophenyl-phenylether	0.400	0.376	6.0	96	0.00
21	Hexachlorobenzene	0.400	0.397	0.8	96	0.00
22	Pentachlorophenol	0.400	0.422	-5.5	71	0.00
23	Phenanthrene	0.400	0.399	0.3	95	0.00
24	Anthracene	0.400	0.363	9.3	92	0.00
25 SURR	Fluoranthene-d10	0.400	0.360	10.0	88	0.00
26	Fluoranthene	0.400	0.375	6.3	89	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	86	0.00
28	Pvrene	0.400	0.417	-4.2	87	0.00
29 S	Terphenyl-d14	0.400	0.391	2.3	87	0.00
30	Benzo(a)anthracene	0.400	0.377	5.8	82	0.00
31	Chrysene	0.400	0.403	-0.8	86	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.246	38.5#	57	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.389	2.8	83	0.00
34 I	Perylene-d12	0.400	0.400	0.0	86	0.00
35	Benzo(b)fluoranthene	0.400	0.366	8.5	83	0.00
36	Benzo(k)fluoranthene	0.400	0.378	5.5	86	0.00
37 C	Benzo(a)pyrene	0.400	0.367	8.3	82	0.00
38	Dibenzo(a,h)anthracene	0.400	0.371	7.3	84	0.00
39	Benzo(a,h,i)perylene	0.400	0.374	6.5	83	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE093019\
Data File : BE100586.D
Acq On : 30 Sep 2019 18:03
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 30 23:07:29 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 27 18:32:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				