

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE100318\
 Data File : BE097712.D
 Acq On : 3 Oct 2018 12:51
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE100318

Quant Time: Oct 03 14:24:55 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 03 12:17:19 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	100	0.00
2	1,4-Dioxane	0.400	0.387	3.3	104	0.00
3	n-Nitrosodimethylamine	0.400	0.345	13.8	91	0.00
4 S	2-Fluorophenol	0.400	0.375	6.3	98	0.00
5 S	Phenol-d6	0.400	0.363	9.3	93	0.00
6	bis(2-Chloroethyl)ether	0.400	0.375	6.3	97	0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	98	0.01
8 S	Nitrobenzene-d5	0.400	0.433	-8.2	114	0.01
9	Naphthalene	0.400	0.389	2.8	98	0.01
10	Hexachlorobutadiene	0.400	0.386	3.5	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.385	3.8	97	0.01
12	2-Methylnaphthalene	0.400	0.392	2.0	100	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.400	0.360	10.0	114	0.00
15 S	2-Fluorobiphenyl	0.400	0.368	8.0	99	0.00
16	Acenaphthylene	0.400	0.364	9.0	97	0.01
17	Acenaphthene	0.400	0.367	8.3	98	0.01
18	Fluorene	0.400	0.370	7.5	99	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	99	0.00
20	4-Bromophenyl-phenylether	0.400	0.374	6.5	99	0.00
21	Hexachlorobenzene	0.400	0.392	2.0	102	0.01
22	Pentachlorophenol	0.400	0.375	6.3	105	0.00
23	Phenanthrene	0.400	0.385	3.8	100	0.01
24	Anthracene	0.400	0.369	7.8	99	0.01
25 SURR	Fluoranthene-d10	0.400	0.366	8.5	97	0.00
26	Fluoranthene	0.400	0.370	7.5	98	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	92	0.00
28	Pvrene	0.400	0.378	5.5	96	0.00
29 S	Terphenyl-d14	0.400	0.379	5.3	97	0.00
30	Benzo(a)anthracene	0.400	0.356	11.0	91	0.00
31	Chrysene	0.400	0.388	3.0	96	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.348	13.0	93	0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.360	10.0	90	0.00
34 I	Perylene-d12	0.400	0.400	0.0	93	0.00
35	Benzo(b)fluoranthene	0.400	0.347	13.3	90	0.00
36	Benzo(k)fluoranthene	0.400	0.365	8.8	94	0.00
37 C	Benzo(a)pyrene	0.400	0.346	13.5	90	0.00
38	Dibenzo(a,h)anthracene	0.400	0.351	12.3	91	0.00
39	Benzo(a,h,i)perylene	0.400	0.363	9.3	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE100318\
Data File : BE097712.D
Acq On : 3 Oct 2018 12:51
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE100318

Quant Time: Oct 03 14:24:55 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 03 12:17:19 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			