

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE100518\
 Data File : BE097771.D
 Acq On : 6 Oct 2018 1:46
 Operator : SJ/JU
 Sample : J5284-14
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 RE-104D1-20181003

Quant Time: Oct 06 04:21:34 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 03 12:17:19 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.91	152	4026	0.40	ng	0.00
7) Naphthalene-d8	10.71	136	16023	0.40	ng	0.01
13) Acenaphthene-d10	14.57	164	8805	0.40	ng	0.01
19) Phenanthrene-d10	17.30	188	25646	0.40	ng	0.00
27) Chrysene-d12	21.49	240	31878	0.40	ng	0.00
34) Perylene-d12	24.04	264	29114	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.48	112	1832	0.19	ng	0.00
5) Phenol-d6	7.05	99	1424	0.10	ng	0.00
8) Nitrobenzene-d5	9.06	82	4210	0.81	ng	0.01
11) 2-Methylnaphthalene-d10	12.31	152	10238	0.40	ng	0.01
14) 2,4,6-Tribromophenol	16.05	330	908	0.57	ng	0.00
15) 2-Fluorobiphenyl	13.19	172	23989	0.62	ng	0.00
25) Fluoranthene-d10	19.34	212	143847	0.44	ng	0.00
29) Terphenyl-d14	19.93	244	41331	0.64	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.39	88	25997	3.241	ng	96
9) Naphthalene	10.76	128	4555	0.028	ng	# 94
32) Bis(2-ethylhexyl)phthalate	21.41	149	6259	0.100	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE100518\
Data File : BE097771.D
Acq On : 6 Oct 2018 1:46
Operator : SJ/JU
Sample : J5284-14
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
RE-104D1-20181003

Quant Time: Oct 06 04:21:34 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 03 12:17:19 2018
Response via : Initial Calibration

