

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE100518\
 Data File : BE097777.D
 Acq On : 6 Oct 2018 5:27
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 06 06:29:17 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 03 12:17:19 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	68	-0.01
2	1,4-Dioxane	0.400	0.376	6.0	69	0.00
3	n-Nitrosodimethylamine	0.400	0.344	14.0	62	0.00
4 S	2-Fluorophenol	0.400	0.407	-1.7	73	-0.01
5 S	Phenol-d6	0.400	0.395	1.3	69	0.00
6	bis(2-Chloroethyl)ether	0.400	0.381	4.8	68	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	72	0.00
8 S	Nitrobenzene-d5	0.400	0.713	-78.2#	138	0.00
9	Naphthalene	0.400	0.388	3.0	72	0.00
10	Hexachlorobutadiene	0.400	0.396	1.0	74	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.387	3.3	72	0.00
12	2-Methylnaphthalene	0.400	0.393	1.8	74	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	78	0.00
14 S	2,4,6-Tribromophenol	0.400	0.507	-26.7#	122	-0.01
15 S	2-Fluorobiphenyl	0.400	0.360	10.0	74	-0.02
16	Acenaphthylene	0.400	0.356	11.0	73	0.00
17	Acenaphthene	0.400	0.376	6.0	77	0.00
18	Fluorene	0.400	0.378	5.5	77	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	76	-0.01
20	4-Bromophenyl-phenylether	0.400	0.376	6.0	76	-0.01
21	Hexachlorobenzene	0.400	0.355	11.3	71	0.00
22	Pentachlorophenol	0.400	0.563	-40.7#	120	-0.01
23	Phenanthrene	0.400	0.380	5.0	75	0.00
24	Anthracene	0.400	0.365	8.8	75	0.00
25 SURR	Fluoranthene-d10	0.400	0.377	5.8	77	0.00
26	Fluoranthene	0.400	0.376	6.0	76	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	70	-0.01
28	Pvrene	0.400	0.395	1.3	77	0.00
29 S	Terphenyl-d14	0.400	0.401	-0.3	79	0.00
30	Benzo(a)anthracene	0.400	0.375	6.3	74	-0.01
31	Chrysene	0.400	0.393	1.8	75	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.468	-17.0	95	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.373	6.8	71	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	75	-0.02
35	Benzo(b)fluoranthene	0.400	0.359	10.3	75	-0.01
36	Benzo(k)fluoranthene	0.400	0.376	6.0	78	0.00
37 C	Benzo(a)pyrene	0.400	0.355	11.3	74	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.341	14.8	71	-0.02
39	Benzo(a,h,i)perylene	0.400	0.357	10.8	73	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE100518\
Data File : BE097777.D
Acq On : 6 Oct 2018 5:27
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 06 06:29:17 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 03 12:17:19 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				