

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE100818\
 Data File : BE097794.D
 Acq On : 8 Oct 2018 14:59
 Operator : SJ/JU
 Sample : PB113550BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB113550BL

Quant Time: Oct 09 06:57:22 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 08 06:37:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.88	152	4528	0.40	ng	-0.01
7) Naphthalene-d8	10.69	136	16960	0.40	ng	0.00
13) Acenaphthene-d10	14.54	164	8000	0.40	ng	-0.02
19) Phenanthrene-d10	17.29	188	18952	0.40	ng	-0.01
27) Chrysene-d12	21.47	240	20653	0.40	ng	-0.01
34) Perylene-d12	24.02	264	23078	0.40	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.47	112	3590	0.34	ng	0.00
5) Phenol-d6	7.04	99	4539	0.28	ng	0.00
8) Nitrobenzene-d5	9.04	82	2995	0.54	ng	0.00
11) 2-Methylnaphthalene-d10	12.30	152	8319	0.31	ng	0.00
14) 2,4,6-Tribromophenol	16.03	330	416	0.29	ng	-0.01
15) 2-Fluorobiphenyl	13.17	172	12552	0.36	ng	-0.02
25) Fluoranthene-d10	19.33	212	72975	0.30	ng	0.00
29) Terphenyl-d14	19.92	244	13047	0.31	ng	-0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE100818\
Data File : BE097794.D
Acq On : 8 Oct 2018 14:59
Operator : SJ/JU
Sample : PB113550BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB113550BL

Quant Time: Oct 09 06:57:22 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 08 06:37:06 2018
Response via : Initial Calibration

