

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE100818\
 Data File : Be097800.d
 Acq On : 8 Oct 2018 19:19
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 09 06:58:40 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 08 06:37:06 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	71	-0.01
2	1,4-Dioxane	0.797	0.755	5.3	72	0.00
3	n-Nitrosodimethylamine	0.725	0.649	10.5	67	0.00
4 S	2-Fluorophenol	0.946	0.890	5.9	70	-0.01
5 S	Phenol-d6	1.450	1.272	12.3	64	-0.01
6	bis(2-Chloroethyl)ether	1.501	1.439	4.1	71	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	75	0.00
8 S	Nitrobenzene-d5	0.130	0.202	-55.4#	124	-0.01
9	Naphthalene	4.031	3.873	3.9	73	-0.01
10	Hexachlorobutadiene	0.214	0.208	2.8	75	-0.01
11 SURR	2-Methylnaphthalene-d10	0.636	0.617	3.0	74	-0.02
12	2-Methylnaphthalene	0.645	0.619	4.0	75	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	80	-0.02
14 S	2,4,6-Tribromophenol	0.072	0.078	-8.3	107	-0.01
15 S	2-Fluorobiphenyl	1.761	1.580	10.3	76	-0.02
16	Acenaphthylene	1.771	1.535	13.3	73	0.00
17	Acenaphthene	1.135	1.055	7.0	79	0.00
18	Fluorene	1.507	1.357	10.0	76	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	78	-0.01
20	4-Bromophenyl-phenylether	0.210	0.198	5.7	79	-0.01
21	Hexachlorobenzene	0.236	0.217	8.1	75	0.00
22	Pentachlorophenol	0.037	0.042	-13.5	99	-0.01
23	Phenanthrene	1.031	0.959	7.0	76	0.00
24	Anthracene	0.890	0.776	12.8	74	-0.01
25 SURR	Fluoranthene-d10	5.063	4.613	8.9	76	-0.01
26	Fluoranthene	1.313	1.199	8.7	76	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	68	-0.01
28	Pvrene	1.016	1.027	-1.1	77	-0.01
29 S	Terphenyl-d14	0.817	0.819	-0.2	77	-0.01
30	Benzo(a)anthracene	1.075	0.977	9.1	69	-0.01
31	Chrysene	1.202	1.163	3.2	72	-0.01
32	Bis(2-ethylhexyl)phthalate	0.787	0.717	8.9	72	-0.01
33	Indeno(1,2,3-cd)pyrene	1.044	0.881	15.6	63	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	68	-0.02
35	Benzo(b)fluoranthene	1.098	0.967	11.9	67	-0.02
36	Benzo(k)fluoranthene	1.217	1.188	2.4	74	-0.02
37 C	Benzo(a)pyrene	0.994	0.845	15.0	65	-0.02
38	Dibenzo(a,h)anthracene	0.979	0.846	13.6	66	-0.03
39	Benzo(a,h,i)perylene	1.024	0.923	9.9	67	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE100818\
Data File : Be097800.d
Acq On : 8 Oct 2018 19:19
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 09 06:58:40 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 08 06:37:06 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		