


Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE100818\
 Data File : Be097801.d
 Acq On : 8 Oct 2018 19:58
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE100318.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Oct 09 11:59:33 2018



AutoFind: Scans 2579, 2580, 2581; Background Corrected with Scan 2572

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	54.1	30675	PASS
68	69	0.00	2	0.7	175	PASS
69	198	0.00	100	46.6	26394	PASS
70	69	0.00	2	0.2	63	PASS
127	198	10	80	46.2	26199	PASS
197	198	0.00	2	0.6	323	PASS
198	198	100	100	100.0	56688	PASS
199	198	5	9	6.8	3829	PASS
275	198	10	60	26.8	15219	PASS
365	198	1	100	3.2	1839	PASS
441	198	0.01	100	13.4	7602	PASS
442	198	50	100	76.7	43505	PASS
443	442	15	24	19.0	8281	PASS