

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE100819\
 Data File : BE100603.D
 Acq On : 8 Oct 2019 11:42
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 09 01:49:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 09 01:42:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.674	0.621	7.9	98	0.00
3	n-Nitrosodimethylamine	0.317	0.291	8.2	114	0.00
4 S	2-Fluorophenol	1.080	0.939	13.1	96	0.00
5 S	Phenol-d6	1.357	1.237	8.8	102	0.00
6	bis(2-Chloroethyl)ether	1.257	1.211	3.7	106	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
8 S	Nitrobenzene-d5	0.250	0.218	12.8	106	0.00
9	Naphthalene	4.126	3.997	3.1	106	0.00
10	Hexachlorobutadiene	0.133	0.130	2.3	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.599	0.572	4.5	106	0.00
12	2-Methylnaphthalene	0.678	0.650	4.1	108	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.091	0.072	20.9	97	0.00
15 S	2-Fluorobiphenyl	1.451	1.419	2.2	108	0.00
16	Acenaphthylene	1.699	1.487	12.5	102	0.00
17	Acenaphthene	1.138	1.061	6.8	107	0.00
18	Fluorene	1.454	1.376	5.4	111	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
20	4-Bromophenyl-phenylether	0.198	0.183	7.6	109	0.00
21	Hexachlorobenzene	0.167	0.164	1.8	111	0.00
22	Pentachlorophenol	0.046	0.022	52.2#	76	0.00
23	Phenanthrene	1.138	1.124	1.2	109	0.00
24	Anthracene	0.979	0.829	15.3	103	0.00
25 SURR	Fluoranthene-d10	4.934	4.494	8.9	109	0.00
26	Fluoranthene	1.401	1.336	4.6	111	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
28	Pvrene	1.025	1.081	-5.5	110	0.00
29 S	Terphenyl-d14	0.884	0.899	-1.7	112	0.00
30	Benzo(a)anthracene	1.118	0.998	10.7	99	0.00
31	Chrysene	1.196	1.181	1.3	105	0.00
32	Bis(2-ethylhexyl)phthalate	2.136	1.170	45.2#	67	0.00
33	Indeno(1,2,3-cd)pyrene	1.409	1.317	6.5	99	0.00
34 I	Perylene-d12	1.000	1.000	0.0	105	0.00
35	Benzo(b)fluoranthene	1.084	0.954	12.0	103	0.00
36	Benzo(k)fluoranthene	1.232	1.165	5.4	106	0.00
37 C	Benzo(a)pyrene	0.995	0.856	14.0	99	0.00
38	Dibenzo(a,h)anthracene	1.047	0.916	12.5	100	0.00
39	Benzo(a,h,i)perylene	1.060	0.951	10.3	100	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE100819\
Data File : BE100603.D
Acq On : 8 Oct 2019 11:42
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 09 01:49:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 09 01:42:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			