

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE100819\
 Data File : BE100614.D
 Acq On : 8 Oct 2019 18:28
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 09 01:46:15 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 09 01:42:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	106	0.00
2	1,4-Dioxane	0.400	0.393	1.8	107	0.00
3	n-Nitrosodimethylamine	0.400	0.435	-8.7	138	-0.01
4 S	2-Fluorophenol	0.400	0.442	-10.5	125	0.00
5 S	Phenol-d6	0.400	0.452	-13.0	130	0.00
6	bis(2-Chloroethyl)ether	0.400	0.401	-0.3	113	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	115	0.00
8 S	Nitrobenzene-d5	0.400	0.409	-2.2	130	0.00
9	Naphthalene	0.400	0.394	1.5	113	0.00
10	Hexachlorobutadiene	0.400	0.375	6.3	107	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.395	1.3	115	-0.02
12	2-Methylnaphthalene	0.400	0.402	-0.5	119	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.400	0.407	-1.7	137	0.00
15 S	2-Fluorobiphenyl	0.400	0.379	5.3	115	0.00
16	Acenaphthylene	0.400	0.387	3.3	124	0.00
17	Acenaphthene	0.400	0.394	1.5	124	0.00
18	Fluorene	0.400	0.387	3.3	124	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	118	0.00
20	4-Bromophenyl-phenylether	0.400	0.380	5.0	121	0.00
21	Hexachlorobenzene	0.400	0.376	6.0	114	0.00
22	Pentachlorophenol	0.400	0.590	-47.5#	176	-0.01
23	Phenanthrene	0.400	0.401	-0.3	120	0.00
24	Anthracene	0.400	0.385	3.8	126	0.00
25 SURR	Fluoranthene-d10	0.400	0.390	2.5	126	0.00
26	Fluoranthene	0.400	0.396	1.0	124	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	114	0.00
28	Pvrene	0.400	0.450	-12.5	127	0.00
29 S	Terphenyl-d14	0.400	0.418	-4.5	124	0.00
30	Benzo(a)anthracene	0.400	0.398	0.5	119	0.00
31	Chrysene	0.400	0.391	2.3	112	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.453	-13.2	150	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.387	3.3	110	0.00
34 I	Perylene-d12	0.400	0.400	0.0	110	0.00
35	Benzo(b)fluoranthene	0.400	0.370	7.5	113	0.00
36	Benzo(k)fluoranthene	0.400	0.363	9.3	107	0.00
37 C	Benzo(a)pyrene	0.400	0.381	4.8	115	0.00
38	Dibenzo(a,h)anthracene	0.400	0.368	8.0	110	0.00
39	Benzo(a,h,i)perylene	0.400	0.371	7.3	108	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE100819\
Data File : BE100614.D
Acq On : 8 Oct 2019 18:28
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 09 01:46:15 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 09 01:42:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				